



facility 755652
facility 755653
project 10243

Isotopic Polonium Case Narrative

COGCC
PW NORM 2017 – 10048

Work Order Number: 1706286

1. This report consists of analytical results and supporting documentation for two water samples received by ALS on 06/13/2017.
2. These samples were prepared according to the current revisions SOP 776 and SOP 711.
3. The samples were analyzed for the presence of Polonium-210 according to the current revision of SOP 714. The analyses were completed on 06/25/2017.
4. The analysis results for these samples are reported in units of pCi/L. The samples were filtered prior to analysis.
5. Results of this analysis are decay-corrected to the sampling date, based on the 138.4 day half-life of Po-210. This decay correction makes no assumptions as to the equilibrium state of Po-210 with the Pb-210 parent nuclide, which has a half-life of 22.3 years.
6. The requested MDC was not met for samples 1706286-1 and -3. These samples were prepared at a reduced aliquot due to suspected matrix interference. These samples were counted for a maximum count time of 1000 minutes and results are reported without further qualification. The results are identified with an “M” qualifier on the final reports.
7. Sample 1706286-3 was initially prepared in batch PL170616-1 on 06/16/2017, at a 50mL aliquot. The sample had a 3% yield after the initial analysis. The sample was re-prepared in batch PL170623-1 at an even further reduced aliquot of 10mL and has a 5.12% yield. All quality control criteria were achieved for the re-preparation. The results of this sample will be reported from batch PL170623-1 at the decision of the client. This sample is identified with an “Y2” flag on the final reports.



8. ALS uses the following convention for reporting significant digits in the TPU and MDC results. The TPU value is rounded to two significant digits. The MDC value is rounded to the same decimal place as the TPU value. In practice, this could result in an MDC reported value of zero for samples with significant activity, including the batch laboratory control sample.
9. No further anomalous situations were encountered during the preparation or analysis of these samples. All remaining quality control criteria were met.

The data contained in the following report have been reviewed and approved by the personnel listed below. In addition, ALS certifies that the analyses reported herein are true, complete and correct within the limits of the methods employed.

Jean Anderson
Jean Anderson

Radiochemistry Primary Data Reviewer

7/24/17
Date

Shibh domy
Radiochemistry Final Data Reviewer

7/26/17
Date

Section 1

CHAIN OF CUSTODY

ALS -- Fort Collins

Sample Number(s) Cross-Reference Table

OrderNum: 1706286

Client Name: COGCC

Client Project Name: PW NORM 2017

Client Project Number: 10048

Client PO Number: CT 2017-3066

Client Sample Number	Lab Sample Number	COC Number	Matrix	Date Collected	Time Collected
755652 Coalview	1706286-1		WATER	13-Jun-17	10:16
755652 Coalview	1706286-2		WATER	13-Jun-17	10:16
755653 Oscar Y	1706286-3		WATER	13-Jun-17	11:36
755653 Oscar Y	1706286-4		WATER	13-Jun-17	11:36

ALS Environmental



225 Commerce Drive, Fort Collins, Colorado 80524
TF: (800) 443-1511 PH: (970) 490-1522

Chain-of-Custody

Turnaround time for samples received after 2 p.m. will be calculated beginning from the next business day.
Turnaround time for samples received Saturday will be calculated beginning from the next business day.

1706286

PROJECT NAME	PW/NORM 2017	SITE ID	EDD FORMAT	45 days	SAMPLER	RC/PAG	PAGE	2 of 3										
PROJECT No.	10048	COGCC					DISPOSAL BY LAB or											
COMPANY NAME	Colorado Oil & Gas Conservation Commission	BILL TO COMPANY	PURCHASE ORDER	CT 2017-3066			PARAMETER/METHOD REQUEST FOR ANALYSIS											
SEND REPORT TO	Peter Gintautas	INVOICE ATTN TO					A total metals SW6010/6020											
ADDRESS	1120 Lincoln St., Suite 801	ADDRESS					B dissolved metals SW6010											
CITY / STATE / ZIP	Denver, CO 80203	CITY / STATE / ZIP					C SW9040A pH											
PHONE	719-679-1326	PHONE					D SM2510B specific conductance											
FAX		FAX					E SM2320C total, bicarbonate and carbonate alkalinity											
E-MAIL	peter.gintautas@state.co.us	E-MAIL					F SM2540C dissolved solids											
							G SM2540D suspended solids											
							H SW9056 anions (Br, Cl, F, SO4)											
							I SAR calculation											
							J											
LAB ID	FIELD ID	MATRIX	SAMPLE DATE	SAMPLE TIME	# OF BOTTLES	PRESERVATIVE	QC	A	B	C	D	E	F	G	H	I	J	SEE NOTES SECTION
755632 Coalview		W	6/13/17	10:16	1	2	X											
755632 Coalview		W	6/13/17	10:16	2	7	X	X	X	X	X	X	X	X	X	X		
755633 Oscar Y		W	6/13/17	11:36	1	2	X											
755633 Oscar Y		W	6/13/17	11:36	2	7	X	X	X	X	X	X	X	X	X	X		
Time Zone (Circle):	MST	Matrix: O = oil S = soil NS = non-soil solid	W = water L = liquid	E = extract F = filter	Form 2029													
NOTES					REPORT LEVEL/QC REQUIRED		RELINQUISHED BY	PETTER GINTAUTAS	DATE	TIME								
6010 total = B, Be, Ca, Cr, Fe, K, Li, Mg, Na, Ni, P, S, Si, V					Summary (Standard QC)		RECEIVED BY		6/13/2017	14:20								
6020 total = Al, Ag, As, Ba, Cd, Co, Cu, Mo, Mn, Na, Pb, Se, Sr, Th, Ti, U, Zn					LEVEL II (Standard QC)		RELINQUISHED BY											
disolved = Ba, Ca, Fe, K, Mg, Na, Si, Sr					LEVEL III (Std QC + forms)		RECEIVED BY											
Dissolved = filter and preserve upon receipt at lab					LEVEL IV (Std QC + forms + raw data)		RELINQUISHED BY											
PRESERVATION KEY	1-HCl 2-HNO3 3-H2SO4 4-NaOH 5-NaOHz/acetate 6-NaHSO4 7-4°C 8-Other						RECEIVED BY											



ALS Environmental

225 Commerce Drive, Fort Collins, Colorado 80524
TF: (800) 443-1511 PH: (970) 490-1511 FAX: (970) 490-1522

Turnaround time for samples received after 2 p.m. will be calculated beginning from the next business day.

Turnaround time for samples received Saturday will be calculated beginning from the next business day.

ALS WORKORDER #

1406786

PROJECT NAME PW NORM 2017

PARAMETER/METHOD REQUEST FOR ANALYSIS

PROJECT NAME	PW NORM 2017	SITE ID	SAMPLER	RC/PAG	PAGE
PROJECT No.	10048	EDD FORMAT	COGCC	DISPOSAL	1 of 3
COMPANY NAME	Colorado Oil & Gas Conservation Commission	PURCHASE ORDER	CT 2017-3066	A	gross alpha/gross beta
SEND REPORT TO	Peter Gintautas	BILL TO COMPANY		B	210Pb
ADDRESS	1120 Lincoln St., Suite 801	INVOICE ATTN TO		C	210Po
CITY / STATE / ZIP	Denver, CO 80203	ADDRESS		D	222Rn
PHONE	719-679-1326	CITY / STATE / ZIP		E	224Ra & 226Ra
FAX		PHONE		F	228Ra
E-MAIL	peter.gintautas@state.co.us	FAX		G	gamma emitters
		E-MAIL		H	*isotopic U
				I	*isotopic Th
				J	
LAB ID	FIELD ID	MATRIX	SAMPLE DATE	# OF BOTTLES	QC
765652 Coalview		W	6/13/17	10/16	3
765652 Coalview		W	6/13/17	10/16	3
755653 Oscar Y		W	6/13/17	11/36	2
755653 Oscar Y		W	6/13/17	11/36	3
Gamma emitters 40K, 137Cs, 212Pb, 212Bi, 214Pb, 224Bi, 226Ra/236U, 228Ac/228Ra, 234mPa, 234Th					
*Time Zone (Circle): MST	Matrix: O = oil S = soil NS = non-soil solid	W = water L = liquid	E = extract F = filter	Form 202e9	SIGNATURE
NOTES					
GAB prepped (coprecip) and counted within 4 days of sampling			REPORT LEVEL / QC REQUIRED	RELINQUISHED BY	PRINTED NAME
224Ra prepped and counted within 4 days of sampling			Summary (Standard QC)	RECEIVED BY	Peter Gintautas
* <input checked="" type="checkbox"/> U if 6020 "total" U >3ug/l			LEVEL II (Standard QC)	RELINQUISHED BY	6/13/2017
* <input checked="" type="checkbox"/> Th only if 6020 "total" Th >3ug/l			LEVEL III (Std QC + forms)	RECEIVED BY	6-13-17
gamma = 40K, 137Cs, 212Pb, 212Bi, 214Pb, 226Ra/235U			LEVEL IV (Std QC + forms + new data)	RELINQUISHED BY	14-26
PRESERVATION KEY			RECEIVED BY		

TIME	DATE
RELINQUISHED BY	
RECEIVED BY	
RELINQUISHED BY	
RECEIVED BY	
RELINQUISHED BY	
RECEIVED BY	



ALS Environmental - Fort Collins
CONDITION OF SAMPLE UPON RECEIPT FORM

Client: COC

Workorder No: 1706286

Project Manager: SS

Initials: JNS Date: 6/13/17

1. Does this project require any special handling in addition to standard ALS procedures?	YES	NO	
2. Are custody seals on shipping containers intact?	<u>NONE</u>	YES	NO
3. Are Custody seals on sample containers intact?	<u>NONE</u>	YES	NO
4. Is there a COC (Chain-of-Custody) present or other representative documents?	<u>YES</u>	NO	
5. Are the COC and bottle labels complete and legible?	<u>YES</u>	NO	
6. Is the COC in agreement with samples received? (IDs, dates, times, no. of samples, no. of containers, matrix, requested analyses, etc.)	<u>YES</u>	NO	
7. Were airbills / shipping documents present and/or removable?	<u>DROP OFF</u>	YES	NO
8. Are all aqueous samples requiring preservation preserved correctly? (excluding volatiles)	N/A	<u>YES</u>	NO
9. Are all aqueous non-preserved samples pH 4-9?	N/A	<u>YES</u>	NO
10. Is there sufficient sample for the requested analyses?	<u>YES</u>	NO	
11. Were all samples placed in the proper containers for the requested analyses?	<u>YES</u>	NO	
12. Are all samples within holding times for the requested analyses?	<u>YES</u>	NO	
13. Were all sample containers received intact? (not broken or leaking, etc.)	<u>YES</u>	NO	
14. Are all samples requiring no headspace (VOC, GRO, RSK/MEE, Rx CN/S, radon) headspace free? Size of bubble: _____ < green pea _____ > green pea	N/A	<u>YES</u>	NO
15. Do any water samples contain sediment? Amount of sediment: _____ dusting <u>X</u> moderate _____ heavy	Amount N/A	<u>YES</u>	NO
16. Were the samples shipped on ice?	<u>YES</u>	NO	
17. Were cooler temperatures measured at 0.1-6.0°C? Cooler #: <u>1</u> <u>2</u> <u>3</u> <u>4</u> Temperature (°C): <u>46</u> <u>46</u> <u>4</u> <u>3.6</u> No. of custody seals on cooler: <u>0</u> <u>0</u> <u>0</u> <u>0</u>	IR gun used*: #2 <u>#4</u> RAD ONLY <u>YES</u>	<u>YES</u>	NO
DOT Survey/Acceptance Information External µR/hr reading: <u>1.0</u> Background µR/hr reading: <u>10</u>			
Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? YES / NO <u>NA</u> (if no, see Form 008.)			

Additional Information: PROVIDE DETAILS BELOW FOR A NO RESPONSE TO ANY QUESTION ABOVE, EXCEPT #1 AND #16.

If applicable, was the client contacted? YES / NO NA Contact: _____ Date/Time: _____

Project Manager Signature / Date: Lilab Denny

*IR Gun #2: Oakton, SN 29922500201-0066

*IR Gun #4: Oakton, SN 2372220101-0002

Section 2

2

SAMPLE RESULTS SUMMARY

Polonium-210 by Alpha Spectroscopy Sample Results Summary

Client Name: COGCC

Client Project Name: PW NORM 2017

Client Project Number: 10048

Laboratory Name: ALS -- Fort Collins

PAL Work Order: 1706286

Page: 1 of 1
Reported on: Wednesday, July 05, 2017
2:54:09 PM

Lab Sample ID	Client Sample ID	Sample Type	Nuclide	Result +/- 2 s TPU	MDC	DL	Units	Matrix	Prep Batch	Date Analyze	Flags
1706286-1	755652 Coalview	Sample	Po-210	0E+00 +/- 3.5E-01	6.6E-01	NA	pCi/l	WATER	PL170616-1	6/21/2017	U,M
1706286-3	755653 Oscar Y	Sample	Po-210	0E+00 +/- 3E+01	5.6E+01	NA	pCi/l	WATER	PL170623-1	6/25/2017	Y2,U,M

Comments:

Data Package ID: PL1706286-1

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

LT - Result is less than Requested MDC, greater than sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

M - The requested MDC was not met.

M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

Abbreviations:

TPU - Total Propagated Uncertainty

MDC - Sample specific Minimum Detectable Concentration

BDL - Below Detection Limit

Section 3

QC RESULTS SUMMARY

3

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 13

Method Blank Results

Lab Name: ALS -- Fort Collins

Work Order Number: 1706286

Client Name: COGCC

ClientProject ID: PW NORM 2017 10048

Lab ID: PL170616-1MB	Sample Matrix: WATER Prep SOP: PAI 711 Rev 10	Prep Batch: PL170616-1 QCBatchID: PL170616-1A Run ID: PL170616-1A Count Time: 1000 minutes	Final Aliquot: 500 ml Result Units: pCi/l File Name: Spectrum #1
	Date Collected: 16-Jun-17 Date Prepared: 16-Jun-17 Date Analyzed: 21-Jun-17		

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Requested MDC	DL	Lab Qualifier
13981-52-7	Po-210	2.1E-02 +/- 2E-02	2.8E-02	5E-01	NA	U

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	9.940E+00	9.3E+00	pCi/l	93.3	30 - 110 %	

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

LT - Result is less than Requested MDC, greater than sample specific MDC.

Abbreviations:

TPU - Total Propagated Uncertainty

MDC - Sample specific Minimum Detectable Concentration

BDL - Below Detection Limit

M - Requested MDC not met.

B - Analyte concentration greater than MDC.

B3 - Analyte concentration greater than MDC but less than Requested MDC.

DL - Decision Level

Data Package ID: PL1706286-1

Date Printed: Wednesday, July 05, 2017

ALS -- Fort Collins

LIMS Version: 6.843

Page 1 of 2

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 13

Method Blank Results

Lab Name: ALS -- Fort Collins

Work Order Number: 1706286

Client Name: COGCC

ClientProject ID: PW NORM 2017 10048

Lab ID: PL170623-1MB	Sample Matrix: WATER Prep SOP: PAI 711 Rev 10	Prep Batch: PL170623-1 QCBatchID: PL170623-1A Run ID: PL170623-1A Count Time: 1000 minutes	Final Aliquot: 500 ml Result Units: pCi/l File Name: Spectrum #1
	Date Collected: 23-Jun-17 Date Prepared: 23-Jun-17 Date Analyzed: 25-Jun-17		

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Requested MDC	DL	Lab Qualifier
13981-52-7	Po-210	6E-03 +/- 1.8E-02	3.4E-02	5E-01	NA	U

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	9.940E+00	9.1E+00	pCi/l	91.7	30 - 110 %	

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

LT - Result is less than Requested MDC, greater than sample specific MDC.

Abbreviations:

TPU - Total Propagated Uncertainty

MDC - Sample specific Minimum Detectable Concentration

BDL - Below Detection Limit

M - Requested MDC not met.

B - Analyte concentration greater than MDC.

B3 - Analyte concentration greater than MDC but less than Requested MDC.

DL - Decision Level

Data Package ID: PL1706286-1

Date Printed: Wednesday, July 05, 2017

ALS -- Fort Collins

LIMS Version: 6.843

Page 2 of 2

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 13

Laboratory Control Sample(s)

Lab Name: ALS -- Fort Collins

Work Order Number: 1706286

Client Name: COGCC

ClientProject ID: PW NORM 2017 10048

Lab ID: PL170616-1LCS	Sample Matrix: WATER Prep SOP: PAI 711 Rev 10	Prep Batch: PL170616-1 QCBatchID: PL170616-1-A Run ID: PL170616-1A Count Time: 480 minutes	Final Aliquot: 500 ml Result Units: pCi/l File Name: Spectrum #1
	Date Collected: 16-Jun-17 Date Prepared: 16-Jun-17 Date Analyzed: 21-Jun-17		

CASNO	Target Nuclide	Results +/- 2s TPU	MDC	Spike Added	% Rec	Control Limits	Lab Qualifier
13981-52-7	Po-210	1.12E+01 +/- 1.7E+00	0E+00	1.070E+01	105	83 - 117	P

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	9.940E+00	8.7E+00	pCi/l	87.5	30 - 110 %	

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

LT - Result is less than Requested MDC, greater than sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

L - LCS Recovery below lower control limit.

H - LCS Recovery above upper control limit.

P - LCS Recovery within control limits.

M - The requested MDC was not met.

M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

Abbreviations:

TPU - Total Propagated Uncertainty

MDC - Minimum Detectable Concentration

Data Package ID: PL1706286-1

Date Printed: Wednesday, July 05, 2017

ALS -- Fort Collins

LIMS Version: 6.843

Page 1 of 2

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 13

Laboratory Control Sample(s)

Lab Name: ALS -- Fort Collins

Work Order Number: 1706286

Client Name: COGCC

ClientProject ID: PW NORM 2017 10048

Lab ID: PL170623-1LCS	Sample Matrix: WATER Prep SOP: PAI 711 Rev 10	Prep Batch: PL170623-1 QCBatchID: PL170623-1-A Run ID: PL170623-1A Count Time: 480 minutes	Final Aliquot: 500 ml Result Units: pCi/l File Name: Spectrum #1
	Date Collected: 23-Jun-17 Date Prepared: 23-Jun-17 Date Analyzed: 25-Jun-17		

CASNO	Target Nuclide	Results +/- 2s TPU	MDC	Spike Added	% Rec	Control Limits	Lab Qualifier
13981-52-7	Po-210	1.1E+01 +/- 1.7E+00	0E+00	1.070E+01	103	83 - 117	P

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	9.940E+00	9.6E+00	pCi/l	96.4	30 - 110 %	

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

LT - Result is less than Requested MDC, greater than sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

L - LCS Recovery below lower control limit.

H - LCS Recovery above upper control limit.

P - LCS Recovery within control limits.

M - The requested MDC was not met.

M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

Abbreviations:

TPU - Total Propagated Uncertainty

MDC - Minimum Detectable Concentration

Data Package ID: PL1706286-1

Date Printed: Wednesday, July 05, 2017

ALS -- Fort Collins

LIMS Version: 6.843

Page 2 of 2

Section 4

INDIVIDUAL SAMPLE RESULTS

4

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 13

Sample Results

Lab Name: ALS -- Fort Collins

Work Order Number: 1706286

Client Name: COGCC

ClientProject ID: PW NORM 2017 10048

Field ID:	755652 Coalview
Lab ID:	1706286-1

Sample Matrix: WATER
Prep SOP: PAI 711 Rev 10
Date Collected: 13-Jun-17
Date Prepared: 16-Jun-17
Date Analyzed: 21-Jun-17

Prep Batch: PL170616-1
QCBatchID: PL170616-1-1
Run ID: PL170616-1A
Count Time: 1000 minutes
Report Basis: Filtered

Final Aliquot: 25.0 ml
Prep Basis: Filtered
Moisture(%): NA
Result Units: pCi/l
File Name: Spectrum #1

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Requested MDC	DL	Lab Qualifier
13981-52-7	Po-210	0E+00 +/- 3.5E-01	6.6E-01	5E-01	NA	U,M

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	1.990E+02	1.67E+02	pCi/l	83.9	30 - 110 %	

Comments: This sample was filtered prior to analysis.

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

LT - Result is less than Requested MDC, greater than sample specific MDC.

M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

M - The requested MDC was not met.

Abbreviations:

TPU - Total Propagated Uncertainty

MDC - Sample specific Minimum Detectable Concentration

BDL - Below Detection Limit

DL - Decision Level

Data Package ID: PL1706286-1

Date Printed: Wednesday, July 05, 2017

ALS -- Fort Collins

LIMS Version: 6.843

Page 1 of 2

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 13

Sample Results

Lab Name: ALS -- Fort Collins

Work Order Number: 1706286

Client Name: COGCC

ClientProject ID: PW NORM 2017 10048

Field ID:	755653 Oscar Y
Lab ID:	1706286-3

Sample Matrix: WATER
Prep SOP: PAI 711 Rev 10
Date Collected: 13-Jun-17
Date Prepared: 23-Jun-17
Date Analyzed: 25-Jun-17

Prep Batch: PL170623-1
QCBatchID: PL170623-1-1
Run ID: PL170623-1A
Count Time: 1000 minutes
Report Basis: Unfiltered

Final Aliquot: 5.00 ml
Prep Basis: Unfiltered
Moisture(%): NA
Result Units: pCi/l
File Name: Spectrum #1

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Requested MDC	DL	Lab Qualifier
13981-52-7	Po-210	0E+00 +/- 3E+01	5.6E+01	5E-01	NA	Y2,U,M

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	9.940E+02	5.1E+01	pCi/l	5.12	30 - 110 %	Y2

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

LT - Result is less than Requested MDC, greater than sample specific MDC.

M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

M - The requested MDC was not met.

Abbreviations:

TPU - Total Propagated Uncertainty

MDC - Sample specific Minimum Detectable Concentration

BDL - Below Detection Limit

DL - Decision Level

Data Package ID: PL1706286-1

Date Printed: Wednesday, July 05, 2017

ALS -- Fort Collins

LIMS Version: 6.843

Page 2 of 2

Section 5

RAW DATA

5

Polonium-210 by Alpha Spectroscopy Raw Data Report

Laboratory Name: ALS -- Fort Collins
 PAI Work Order: 1706286

Prep SOP: PAI 711
 Analytical SOP: PAI 714

Reported on: Monday, July 03, 2017
 2:29:56 PM

Sample ID QC Type	Nuclide Type	Sample Date/Time	Prep Batch QCBatchID	Ingrowth Date/Time	Decay Date/Time	Matrix %Moist.	Samp Alq Analy Alq	Inst ID Det ID	AnRunID File Name	Count Date/Time	Net Cnts Bkg Cnts	BaseEff Bkg/min	CntDur/min)	Activity +/- 2 s TPU	MDC	ReportUnits ReportBasis	DER RPD	&Spk/ Recov Flags				
1706286-1	Po-209	6/13/2017	PL170616-1	NA	6/19/2017	WATER	50 mL	25 mL	AlphaSpec2	PL170616-1A Spectrum #1	6/21/2017 10:46 AM	4,000	1000	31.73%	1E+00	pCi/l	NA	NA				
SMP	Tracer	10:16:00 AM	PL170616-1	NA	3:50:00 PM	NA	6/19/2017	WATER	50 mL	25 mL	AlphaSpec2	PL170616-1A Spectrum #1	6/21/2017 10:46 AM	2,000	1000	31.73%	1.1E+01	pCi/l	NA	U,M		
1706286-1	Po-210	6/13/2017	PL170616-1	NA	3:50:00 PM	NA	6/19/2017	WATER	50 mL	25 mL	AlphaSpec2	PL170616-1A Spectrum #1	6/21/2017 10:46 AM	2,000	1000	31.73%	3.5E+01	Filter	NA	NA		
SMP	Trg. Analyte	10:16:00 AM	PL170616-1	NA	4:08:00 PM	NA	6/24/2017	WATER	10 mL	AlphaSpec2	PL170623-1A Spectrum #1	6/25/2017 10:42 AM	3,000	1000	30.78%	1.1E+01	pCi/l	NA	Y2			
1706286-3	Po-209	6/13/2017	PL170623-1	NA	6/24/2017	WATER	5 mL	103	AlphaSpec2	PL170623-1A Spectrum #1	6/25/2017 10:42 AM	2,000	1000	30.78%	1.1E+01	pCi/l	NA	U,M				
SMP	Tracer	11:36:00 AM	PL170623-1	NA	4:08:00 PM	NA	6/19/2017	WATER	5 mL	103	AlphaSpec2	PL170623-1A Spectrum #1	6/25/2017 10:42 AM	2,000	1000	30.78%	3E+01	Unfilter	NA	Y2,U,M		
PL170616-1	Po-210	6/13/2017	PL170623-1	NA	4:08:00 PM	NA	6/19/2017	WATER	1000 mL	AlphaSpec2	PL170616-1A Spectrum #1	6/21/2017 10:46 AM	3,000	1000	32.54%	1.3E+00	pCi/l	NA	NA			
MB	Tracer	2:50:26 PM	PL170616-1	NA	3:50:00 PM	NA	6/19/2017	WATER	1000 mL	AlphaSpec2	PL170616-1A Spectrum #1	6/21/2017 10:46 AM	2,000	1000	32.54%	2.1E-02	pCi/l	NA	NA			
PL170616-1	Po-210	6/16/2017	PL170616-1	NA	3:50:00 PM	NA	6/19/2017	WATER	500 mL	113	AlphaSpec2	PL170616-1A Spectrum #1	6/21/2017 10:46 AM	2,000	1000	32.54%	2E-02	Unfilter	NA	NA		
PL170616-1	Po-209	6/16/2017	PL170616-1	NA	3:50:00 PM	NA	6/19/2017	WATER	1000 mL	AlphaSpec2	PL170616-1A Spectrum #1	6/21/2017 10:47 AM	1,000	1000	32.54%	3.5E+00	pCi/l	NA	NA			
LCS	Tracer	2:50:26 PM	PL170616-1	NA	3:50:00 PM	NA	6/19/2017	WATER	1000 mL	AlphaSpec2	PL170616-1A Spectrum #1	6/21/2017 10:47 AM	1,000	1000	32.54%	3.5E+00	Unfilter	NA	NA			
PL170616-1	Po-210	6/16/2017	PL170616-1	NA	3:50:00 PM	NA	6/19/2017	WATER	1000 mL	AlphaSpec2	PL170616-1A Spectrum #1	6/21/2017 10:47 AM	1,000	1000	32.54%	4.8E-02	pCi/l	NA	NA			
LCS	Trg. Analyte	2:50:26 PM	PL170616-1	NA	3:50:00 PM	NA	6/19/2017	WATER	1000 mL	AlphaSpec2	PL170616-1A Spectrum #1	6/21/2017 10:47 AM	1,000	1000	32.54%	4.8E-02	Unfilter	NA	NA			
PL170623-1	Po-209	6/23/2017	PL170623-1	NA	6/24/2017	WATER	1000 mL	AlphaSpec2	PL170623-1A Spectrum #1	6/25/2017 10:42 AM	3,000	1000	31.90%	480	8.7E+00	0E+00	pCi/l	NA	NA			
MB	Tracer	11:13:20 AM	PL170623-1	NA	4:08:00 PM	NA	500 mL	112	AlphaSpec2	PL170623-1A Spectrum #1	6/25/2017 10:42 AM	1,000	1000	31.90%	1.3E+00	Unfilter	NA	NA				
PL170623-1	Po-210	6/23/2017	PL170623-1	NA	4:08:00 PM	NA	6/24/2017	WATER	1000 mL	AlphaSpec2	PL170623-1A Spectrum #1	6/25/2017 10:42 AM	2,000	1000	31.90%	1.12E+01	pCi/l	NA	105			
MB	Trg. Analyte	11:13:20 AM	PL170623-1	NA	4:08:00 PM	NA	6/24/2017	WATER	500 mL	112	AlphaSpec2	PL170623-1A Spectrum #1	6/25/2017 10:42 AM	3,000	1000	31.90%	1.7E+00	Unfilter	NA	P		
PL170623-1	Po-209	6/23/2017	PL170623-1	NA	6/24/2017	WATER	1000 mL	AlphaSpec2	PL170623-1A Spectrum #1	6/25/2017 10:42 AM	3,000	1000	31.90%	9.1E+00	0E+00	0E+00	pCi/l	NA	NA			
MB	Tracer	11:13:20 AM	PL170623-1	NA	4:08:00 PM	NA	6/24/2017	WATER	500 mL	112	AlphaSpec2	PL170623-1A Spectrum #1	6/25/2017 10:42 AM	1,000	1000	31.90%	1.3E+00	Unfilter	NA	NA		
PL170623-1	Po-210	6/23/2017	PL170623-1	NA	4:08:00 PM	NA	6/19/2017	WATER	1000 mL	AlphaSpec2	PL170623-1A Spectrum #1	6/25/2017 10:42 AM	3,000	1000	31.90%	6E-03	3.4E-02	pCi/l	NA	NA		
MB	Trg. Analyte	11:13:20 AM	PL170623-1	NA	4:08:00 PM	NA	6/19/2017	WATER	1000 mL	AlphaSpec2	PL170623-1A Spectrum #1	6/25/2017 10:43 AM	3,000	1000	31.90%	6.6E-03	3.4E-02	pCi/l	NA	NA		
PL170623-1	Po-209	6/23/2017	PL170623-1	NA	4:08:00 PM	NA	6/19/2017	WATER	500 mL	113	AlphaSpec2	PL170623-1A Spectrum #1	6/25/2017 10:43 AM	3,000	1000	32.54%	480	9.6E+00	0E+00	Unfilter	NA	NA
LCS	Tracer	11:13:20 AM	PL170623-1	NA	4:08:00 PM	NA	6/19/2017	WATER	500 mL	113	AlphaSpec2	PL170623-1A Spectrum #1	6/25/2017 10:43 AM	3,000	1000	32.54%	1.4E+00	Unfilter	NA	NA		

Comments:

Data Package ID: PL17062386-1

Qualifiers/Flags:

U - Result is less than the sample specific MDC.
 Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.

Y2 - Chemical Yield outside default limits.

W - DER is greater than Warning Limit of 1.42

D - DER is greater than Control Limit of 2.13

+ - Duplicate RPD not within limits.

LT - Result is less than Request MDC, greater than sample specific MDC

* - Aliquot Basis is 'As Received' while the Report Basis is 'Dry Weight'

- Aliquot Basis is 'Dry Weight' while the Report Basis is 'As Received'

Notes:
 1) The Tracer results are not yield corrected (i.e. activity measured not activity added).
 2) Where sample time is not available, 12:00 PM (Mountain) is used for decay correction.

Abbreviations:

L - LCS Recovery below lower control limit.

H - LCS Recovery above upper control limit.

P - LCS, Matrix Spike Recovery within control limits

N - Matrix Spike Recovery outside control limits

NC - Not Calculated for duplicate results less than 5 times MDC

B - Analyte concentration greater than MDC

BDL - Below Detection Limit

Date Printed: Wednesday, July 05, 2017
 19 71

ALS -- Fort Collins

LIMS Version: 6.843

Polonium-210 by Alpha Spectroscopy Raw Data Report

Laboratory Name: ALS -- Fort Collins

PAI Work Order: 1706286

Prep SOP: PAI 711

Analytical SOP: PAI 714

Reported on: Monday, July 03, 2017

2:41:14 PM

Sample ID QC Type	Nuclide Type	Sample Date/Time	Prep Batch QCBatchID	Ingrowth Date / Time	Decay Date/Time	Matrix %Moist.	Samp Aliq Analy Aliq	Inst ID Det ID	AnRrunID File Name	Count Date/Time	Net Cnts Bkg Cnts	BaseEff Bkg/min)	CntDur(min) Yield	Activity +/- 2 s TPU	MDC	ReportUnits ReportBasis	DER RPD	&Spk. Recov Flags
PL170623-1	Po-210	6/23/2017	PL170623-1	NA	6/24/2017	WATER	1000 mL	AlphaSpec2	PL170623-1A	6/25/2017	1839.040	32.54%	480	1.1E+01	0E+00	pCi/l	NA	103
LCS	Ttg. Analyte	11:13:20 AM	PL170623-1-1	NA	4:08:00 PM	NA	500 mL	113	Spectrum #1	10:43 AM	2,000	1000	96.4%	1.7E+00	Unfiltered	NA	P	

Comments:

Data Package ID: PL1706286-1

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.
- Y2 - Chemical Yield outside default limits.
- W - DER is greater than Warning Limit of 1.42
- D - DER is greater than Control Limit of 2.13
- + - Duplicate RPD not within limits.
- LT - Result is less than Request MDC, greater than sample specific MDC
- * - Aliquot Basis is 'As Received' while the Report Basis is 'Dry Weight'
- # - Aliquot Basis is 'Dry Weight' while the Report Basis is 'As Received'

Notes:

- 1) The Tracer results are not yield corrected (i.e. activity measured not activity added).
- 2) Where sample time is not available, 12:00 PM (Mountain) is used for decay correction.

Abbreviations:

- L - LCS Recovery below/above upper control limit.
- H - LCS Recovery above upper control limit.
- P - LCS, Matrix Spike Recovery within control limits.
- N - Matrix Spike Recovery outside control limits
- NC - Not Calculated for duplicate results less than 5 times MDC
- B - Analyte concentration greater than MDC.
- BDL - Below Detection Limit

Date Printed: Wednesday, July 05, 2017

ALS -- Fort Collins

LIMS Version: 6.843

Page 2 of 2

Analyst: ORTEC

Sample: 1706286-1
 Spectrum #1 Analysis #1
 :
 Sample Collection Date:
 Comment:

Type: Sample

Sample

Sample Volume : 0.03
 Sample Units: L
 First Stage Dilution: N/A
 Aliquot: N/A Aliquot Fraction: N/A
 Dilution 2: N/A
 Lab Preparation:

Batch Name: PL170616-1_B

Batch

Client Name: Undefined
 Client Contact:

Description:

Tracer


Tracer Name: 1045.4243.07 Po-209
 Tracer Activity: 22.23 DPM / mL x (Vol.) 0.50 mL = 11.11 DPM
 Tracer Ref. Date: 6/8/2016 10:00:00AM

Acquisition

Tracer Nuclide: Po-209
 Tracer Recovery: 84.14%

Detector: 105, SN: 5505338, ID: 105
 Acquisition Start Date: 6/21/2017 10:46:42AM
 Live Time: 1,000.00 min.
 Real Time: 1,000.00 min.
 Background Date: 6/20/2017 12:43:21PM
 Bkgd Info: Sample: B170620105; Det: 105; Spectrum #1; 6/20/2017
 12:43:21 PM; Live Time: 1000.000(min.); ID: 105

Energy Calibration: C170620105A
 Efficiency Calibration: C170620105A
 Calibration Date: 6/20/2017 11:38:25AM
 Energy Cal: Gain = 9.9003 keV / Ch
 Offset = 3,021.28 keV
 Quadratic = 0.0000 keV / Ch²
 Efficiency: 31.73% +/- 1.31% TPU(2 sigma)



General Analysis

Analysis Method: ROI Analysis, Set Name = Po-210

Decay Correction: 6/21/2017 10:45:40AM

Nuclide Library: Polonium

MDA Constants: K_α = 1.64, K_β = 1.64

MDA Source: Background

Nuclide Summary (ROI)

Nuclide	Peak Energy keV	ROI Start keV	ROI End keV	FWHM keV	B.R. %	Gross Counts	Bkgd Counts	Net Counts	Activity pCi/L	1.00Sigma TPU pCi/L	Critical Level pCi/L	MDA pCi/L
Po-209	4882.533	4595.425	4961.735	62.7	99.7	2,943.00	4.0000	2939.00	1.685E+002	4.670E+000	3.171E-001	8.185E-001
Po-210	5298.345	5031.037	5427.048	59.0	100.0	2.00	2.0000	0.00	0.000E+000	5.399E+000	2.220E-001	6.267E-001

JP

Analyst: user

8:42:30AM 6/26/2017

Sample: 1706286-3
 Spectrum #1 Analysis #1
 :
 Sample Collection Date:
 Comment:

Sample

Sample Volume : 0.00
 Sample Units: L
 First Stage Dilution: N/A
 Aliquot: N/A Aliquot Fraction: N/A
 Dilution 2: N/A
 Lab Preparation:

Batch Name: PL170623-1_A

Batch

Client Name: Undefined
 Client Contact:

Description:

Tracer


Tracer Name: 1045.4243.07 Po-209
 Tracer Activity: 22.23 DPM / mL x (Vol.) 0.50 mL = 11.11 DPM
 Tracer Ref. Date: 6/8/2016 10:00:00AM

Acquisition

Tracer Nuclide: Po-209
 Tracer Recovery: 5.14%

Detector: 103, SN: 5505336, ID: 103
 Acquisition Start Date: 6/25/2017 10:42:14AM
 Live Time: 1,000.00 min.
 Real Time: 1,000.00 min.
 Background Date: 6/20/2017 12:43:21PM
 Bkgd Info: Sample: B170620103; Det: 103; Spectrum #1; 6/20/2017
 12:43:21 PM; Live Time: 1000.000(min.); ID: 103

Energy Calibration: C170620103
 Efficiency Calibration: C170620103
 Calibration Date: 6/20/2017 11:38:19AM
 Energy Cal: Gain = 9.9003 keV / Ch
 Offset = 3,021.28 keV
 Quadratic = 0.0000 keV / Ch²
 Efficiency: 30.78% +/- 1.22% TPU(2 sigma)

**General Analysis**

Analysis Method: ROI Analysis, Set Name = Po-210

Decay Correction: 6/25/2017 10:41:12AM

Nuclide Library: Polonium

MDA Constants: K α = 1.64, K β = 1.64

MDA Source: Background

Nuclide Summary (ROI)

Nuclide	Peak Energy keV	ROI Start keV	ROI End keV	FWHM keV	B.R. %	Gross Counts	Bkgd Counts	Net Counts	Activity pCi/L	1.00Sigma TPU pCi/L	Critical Level pCi/L	MDA pCi/L
Po-209	4882.533	4595.425	4961.735	85.6	99.7	177.00	3.0000	174.00	5.143E+001	4.095E+000	2.319E+001	6.195E+001
Po-210	5298.345	5031.037	5427.048	57.6	100.0	2.00	2.0000	0.00	0.000E+000	2.280E+003	1.875E+001	5.293E+001

JN

Sample: PL170616-1MB
 Spectrum #1 Analysis #1
 :
 Sample Collection Date:
 Comment:

Type: Sample

Batch Name: PL170616-1_B
 :
 Description:

Tracer Name: 1045.4243.07 Po-209
 Tracer Activity: 22.23 DPM / mL x (Vol.) 0.50 mL = 11.11 DPM
 Tracer Ref. Date: 6/8/2016 10:00:00AM

Detector: 113, SN: 5505346, ID: 113
 Acquisition Start Date: 6/21/2017 10:46:43AM
 Live Time: 1,000.00 min.
 Real Time: 1,000.06 min.
 Background Date: 6/20/2017 12:44:10PM
 Bkgd Info: Sample: B170620113; Det: 113; Spectrum #1; 6/20/2017
 12:44:10 PM; Live Time: 1000.000(min.); ID: 113

Sample

Sample Volume : 0.03 Sample Units: L
 First Stage Dilution: N/A
 Aliquot: N/A Aliquot Fraction: N/A
 Dilution 2: N/A
 Lab Preparation:

Batch


Client Name: Undefined
 Client Contact:

Tracer

Tracer Nuclide: Po-209
 Tracer Recovery: 93.56%

Acquisition

Energy Calibration: C170620113
 Efficiency Calibration: C170620113
 Calibration Date: 6/20/2017 11:38:44AM
 Energy Cal: Gain = 9.9003 keV / Ch
 Offset = 3,021.28 keV
 Quadratic = 0.0000 keV / Ch²
 Efficiency: 32.54% +/- 1.24% TPU(2 sigma)



General Analysis

Analysis Method: ROI Analysis, Set Name = Po-210

Decay Correction: 6/21/2017 10:45:40AM

MDA Constants: K_α = 1.64, K_β = 1.64

Nuclide Library: Polonium

MDA Source: Background

Nuclide Summary (ROI)

Nuclide	Peak Energy keV	ROI Start keV	ROI End keV	FWHM keV	B.R. %	Gross Counts	Bkgd Counts	Net Counts	Activity pCi/L	1.00Sigma TPU pCi/L	Critical Level pCi/L	MDA pCi/L
Po-209	4882.533	4595.425	4961.735	62.6	99.7	3,355.00	3.0000	3352.00	1.874E+002	4.831E+000	2.407E-001	6.432E-001
Po-210	5298.345	5031.037	5427.048	37.4	100.0	9.00	2.0000	7.00	4.142E-001	1.978E-001	1.947E-001	5.495E-001

[Handwritten Signature]

Sample: PL170616-1LCS
Spectrum #1 Analysis #1

:
Sample Collection Date:
Comment:

Batch Name: PL170616-1_C

:
Description:

Tracer Name: 1045.4243.07 Po-209

Tracer Activity: 22.23 DPM / mL x (Vol.) 0.50 mL = 11.11 DPM

Tracer Ref. Date: 6/8/2016 10:00:00AM

Detector: 114, SN: 5505347, ID: 114

Acquisition Start Date: 6/21/2017 10:47:20AM

Live Time: 480.00 min.

Real Time: 480.00 min.

Background Date: 6/20/2017 12:44:10PM

Bkgd Info: Sample: B170620114; Det: 114; Spectrum #1; 6/20/2017 12:44:10 PM; Live Time: 1000.000(min.); ID: 114

Sample

Sample Volume : 0.50 Sample Units: L
First Stage Dilution: N/A
Aliquot: N/A Aliquot Fraction: N/A
Dilution 2: N/A
Lab Preparation:

Batch


Client Name: Undefined
Client Contact:

Tracer

Tracer Nuclide: Po-209
Tracer Recovery: 87.69%

Acquisition

Energy Calibration: C170620114
Efficiency Calibration: C170620114
Calibration Date: 6/20/2017 11:38:46AM
Energy Cal: Gain = 9.9003 keV / Ch
Offset = 3,021.28 keV
Quadratic = 0.0000 keV / Ch²
Efficiency: 33.42% +/- 2.06% TPU(2 sigma)



General Analysis

Analysis Method: ROI Analysis, Set Name = Po-210

Decay Correction: 6/21/2017 10:46:58AM

MDA Constants: K_α = 1.64, K_β = 1.64

Nuclide Library: Polonium

MDA Source: Background

Nuclide Summary (ROI)

Nuclide	Peak Energy keV	ROI Start keV	ROI End keV	FWHM keV	B.R. %	Gross Counts	Bkgd Counts	Net Counts	Activity pCi/L	1.00Sigma TPU pCi/L	Critical Level pCi/L	MDA pCi/L
Po-209	4882.533	4595.425	4961.735	61.4	99.7	1,549.00	0.4800	1548.52	8.781E+000	3.505E-001	8.966E-003	3.543E-002
Po-210	5298.345	5031.037	5427.048	60.5	100.0	1,742.00	0.4800	1741.52	1.115E+001	8.358E-001	8.880E-003	3.509E-002

[Handwritten Signature]

Analyst: user

Sample: PL170623-1MB
 Spectrum #1 Analysis #1
 :
 Sample Collection Date:
 Comment:

Type: Sample

Sample

Sample Volume : 0.50 Sample Units: L
 First Stage Dilution: N/A
 Aliquot: N/A Aliquot Fraction: N/A
 Dilution 2: N/A
 Lab Preparation:

Batch Name: PL170623-1_A

Batch

Client Name: Undefined
 Client Contact:

Description:

Tracer


Tracer Name: 1045.4243.07 Po-209
 Tracer Activity: 22.23 DPM / mL x (Vol.) 0.50 mL = 11.11 DPM
 Tracer Ref. Date: 6/8/2016 10:00:00AM

Acquisition

Tracer Nuclide: Po-209
 Tracer Recovery: 91.91%

Detector: 112, SN: 5505345, ID: 112
 Acquisition Start Date: 6/25/2017 10:42:15AM
 Live Time: 1,000.00 min.
 Real Time: 1,000.00 min.
 Background Date: 6/20/2017 12:44:10PM
 Bkgd Info: Sample: B170620112; Det: 112; Spectrum #1; 6/20/2017
 12:44:10 PM; Live Time: 1000.000(min.); ID: 112

Energy Calibration: C170620112
 Efficiency Calibration: C170620112
 Calibration Date: 6/20/2017 11:38:41AM
 Energy Cal: Gain = 9.9003 keV / Ch
 Offset = 3,021.28 keV
 Quadratic = 0.0000 keV / Ch²
 Efficiency: 31.90% +/- 1.32% TPU(2 sigma)

**General Analysis**

Analysis Method: ROI Analysis, Set Name = Po-210

Decay Correction: 6/25/2017 10:41:12AM

MDA Constants: K_α = 1.64, K_β = 1.64

Nuclide Library: Polonium

MDA Source: Background

Nuclide Summary (ROI)

Nuclide	Peak Energy keV	ROI Start keV	ROI End keV	FWHM keV	B.R. %	Gross Counts	Bkgd Counts	Net Counts	Activity pCi/L	1.00Sigma TPU pCi/L	Critical Level pCi/L	MDA pCi/L
Po-209	4882.533	4595.425	4961.735	64.1	99.7	3,230.00	3.0000	3,227.00	9.203E+000	2.497E-001	1.250E-002	3.340E-002
Po-210	5298.345	5031.037	5427.048	58.2	100.0	5.00	3.0000	2.00	6.146E-003	8.700E-003	1.238E-002	3.308E-002

JL

Analyst: user

8:43:33AM

6/26/2017

Sample: PL170623-1LCS
 Spectrum #1 Analysis #1
 :
 Sample Collection Date:
 Comment:

Sample

Sample Volume : 0.50

Sample Units: L

First Stage Dilution: N/A

Aliquot: N/A Aliquot Fraction: N/A

Dilution 2: N/A

Lab Preparation:

Batch Name: PL170623-1_B
 :
 Description:

Batch

Client Name: Undefined

Client Contact:

Tracer Name: 1045.4243.07 Po-209
 Tracer Activity: 22.23 DPM / mL x (Vol.) 0.50 mL = 11.11 DPM
 Tracer Ref. Date: 6/8/2016 10:00:00AM
 Detector: 113, SN: 5505346, ID: 113
 Acquisition Start Date: 6/25/2017 10:43:00AM
 Live Time: 480.00 min.
 Real Time: 480.00 min.
 Background Date: 6/20/2017 12:44:10PM
 Bkgd Info: Sample: B170620113; Det: 113; Spectrum #1; 6/20/2017
 12:44:10 PM; Live Time: 1000.000(min.); ID: 113


Tracer

Tracer Nuclide: Po-209

Tracer Recovery: 96.63%

Acquisition

Energy Calibration: C170620113
 Efficiency Calibration: C170620113
 Calibration Date: 6/20/2017 11:38:44AM
 Energy Cal: Gain = 9.9003 keV / Ch
 Offset = 3,021.28 keV
 Quadratic = 0.0000 keV / Ch²
 Efficiency: 32.54% +/- 1.24% TPU(2 sigma)

**General Analysis**

Analysis Method: ROI Analysis, Set Name = Po-210

Decay Correction: 6/25/2017 10:42:31AM

MDA Constants: K_α = 1.64, K_β = 1.64

Nuclide Library: Polonium

MDA Source: Background

Nuclide Summary (ROI)

Nuclide	Peak Energy	ROI Start keV	ROI End keV	FWHM keV	B.R. %	Gross Counts	Bkgd Counts	Net Counts	Activity pCi/L	1.00Sigma TPU pCi/L	Critical Level pCi/L	MDA pCi/L
Po-209	4882.533	4595.425	4961.735	64.7	99.7	1,663.00	1.4400	1661.56	9.676E+000	3.011E-001	1.447E-002	4.525E-002
Po-210	5298.345	5031.037	5427.048	64.2	100.0	1,840.00	0.9600	1839.04	1.098E+001	7.263E-001	1.170E-002	3.956E-002

ALS

Alpha Spectrometer Instrument Run Log

SOP 714;FORM 74678.xls (01/20/07)

Date: 6/21/17

Detector	Batch ID	Sample ID	Iso/Matrix	Duration	Initial
03	PL170621-1-B	17062771-1	Ra/W	1000	50
04				-2	
05		17062866-1			
07				-3	
08		17062999-1			
09				-2	
10		17063299-1			
11				-2	
12	C	1706374-2	480		
13	B	PL170616-1MB	1000		
14	C	LCJ	480		
01	RAS170621-1-A	RAS170621-1MB	Ra/W	1000	50
02					
03				-2	
04		1706421-1	LCJ		
05				-2	
06		1706453-1			
07	6/23/17			-2	
08	6/23/17				
09		1706423-1			
10		1706426-1			
11				-3	
12				-3D	
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					
23					
24					
25					
26					
27					
28					
29					
30					
31					
32					
33					
34					
35					
36					
37					
38					
39					
40					
41					
42					
43					
44					
45					

Detector	Batch ID	Sample ID	Iso/Matrix	Duration	Initial
01	RAS170622-1A	RAS170622-1B	Ra/W	1000	50
02				-2	
03					
04				-3	
05					
06				-4	
07					
08				-5	
09					
10				-4	
11					
12				-2	
13				-3	
14				-4	
15				-5	
16					
17					
18					
19					
20					
21					
22					
23					
24					
25					
26					
27					
28					
29					
30					
31					
32					
33					
34					
35					
36					
37					
38					
39					
40					
41					
42					
43					
44					
45					

Notes:

Reviewed by: *J.P.*
 Date: 6/23/17

ALS

Alpha Spectrometer Instrument Run Log

Detector	Batch ID	Sample ID	Iso/Matrix	Duration	Initial
91	TAS170620-3-B	1706326-6	Th/W	1000	V0
92		-7			
93	1706343-2				
94		-3			
95		-4			
96		-5			
97		-6			
98		-7			
99		-8			
90		-9			
91		1706380-2			
93		-3			
94		-4			
95		-5			
117		AS170620-3mB			
118		LC5			
103	PL170623-1-A	1706286-3	Po/W	1000	V1
104					
105		-3			
107	B	1706421-1			
108		480			
109	A	1706423-1			
110		1000			
111		-3			

Notes:

Detector	Batch ID	Sample ID	Iso/Matrix	Duration	Initial
112	PL170623-1-A	PL170623-1mB	Po/W	1000	V1
113		B			
81	RAS170641-B	RAS170641mB	Rn/W	300	JP
82					
83		LC5			
84		1706271-1			
85		-2			
86		1706278-2			
87		-3			
88		-4			
89		-5			
90		-6			
91		-7			
92		-8			
93		-9			
94		1706286-1			
95		-10			
96		-11			
97		-12			
98		-13			
99		-14			
100		-15			
101		-16			
102		-17			
103		-18			
104		-19			
105		-20			
106		-21			
107		-22			
108		-23			
109		-24			
110		-25			
111		-26			

Reviewed by: *[Signature]*
Date: 6/24/17

471177

Section 6

QUALITY ASSURANCE SUMMARY REPORTS

6

No *NON-COMFORMANCE REPORTS* or *QUALITY ASSURANCE SUMMARY SHEETS* are included in this data package.

Section 7

LABORATORY BENCH SHEETS

7

ALS -- Fort Collins

Radiochemistry Instrument Worksheet

Prep Batch: PL170616-1

Prep Procedure: Po210

US6S LCS → 480m¹
C6,CC, MB → 1000 Mn

Prep Num	LabID	QC Type	Init Alq	Fin Alq	Units	Report Units	Cnt 1 File	Cnt 1 Pos Chk By	Cnt 2 File	Cnt 2 Pos Chk By	Cnt 3 File	Cnt 3 Inst/Det	Cnt 3 Pos Chk By	Notes
7/10	1	1706271-1	SMP	50	25	ml	pCi/l	62711	103	JP	62711		62711	
	1	1706271-2	SMP	50	25	ml	pCi/l	62712	104	JP	62712		62712	
	1	1706270-2	SMP	187.879	93.939	ml	pCi/l	62782	65	JP	62782		62782	
	1	1706278-5	SMP	188	94	ml	pCi/l	62785	67		62785		62785	
	1	1706278-6	SMP	187.817	93.909	ml	pCi/l	62786	69		62786		62786	
	1	1706286-1	SMP	50	25	ml	pCi/l	62861	105	106	62861		62861	
	1	1706286-3	SMP	50	25	ml	pCi/l	62863	107	108	62863		62863	
	1	1706299-1	SMP	50	25	ml	pCi/l	62991	108		62991		62991	
	1	1706299-2	SMP	50	25	ml	pCi/l	62992	109		62992		62992	
	1	1706306-1	SMP	187.879	93.939	ml	pCi/l	63061	70	JP	63061		63061	
	1	1706306-11	SMP	187.94	93.97	ml	pCi/l	630611	72		630611		630611	
	1	1706306-12	SMP	188	94	ml	pCi/l	630612	74		630612		630612	
	1	1706329-1	SMP	50	25	ml	pCi/l	63291	110	JP	63291		63291	
	1	1706329-2	SMP	50	25	ml	pCi/l	63292	111		63292		63292	
	1	1706342-3	SMP	188	94	ml	pCi/l	63423	75	JP	63423		63423	
	1	1706342-3	DUP	188	94	ml	pCi/l	63423D	77		63423D		63423D	
	1	1706342-7	SMP	186.997	93.478	ml	pCi/l	63427	75	78	63427		63427	
	1	1706342-8	SMP	187.5	93.75	ml	pCi/l	63428	79		63428		63428	
	1	1706374-1	SMP	188	94	ml	pCi/l	63741	80		63741		63741	
	1	1706374-2	SMP	187.817	93.909	ml	pCi/l	63742	112	JP	63742		63742	
	1	PL170616-1	MB	1000	500	ml	pCi/l	170611B	113		170611B		170611B	
	1	PL170616-1	LCS	1000	500	ml	pCi/l	170611L	114		170611L		170611L	

JN/NA

Soln #	Nuclide	SolnID	Exp Date	Prep Conc	Units	Prep Date	Aliquot Units	Pipet ID
T1	Po-209	1045.4243.07	22.074	DPM/ml	06/16/17	1 ml	AW026	

Analytical QASS / NCR? Y / N								
Tracer/Carrier Solution Information								
Soln #	Nuclide	SolnID	Exp Date	Prep Conc	Units	Prep Date	Aliquot Units	Pipet ID
S1	Pt-210		899.4095.66	94.877	DPM/ml	06/16/17	0.25 ml	AW016
S1	Po-210		899.4095.66	94.877	DPM/ml	06/16/17	0.25 ml	AW016

Sample Barcodes

of 7

Page 1 of 3 Po210 Instrument Sheet
Date Printed: 6/20/2017 7:52

ALS -- Fort Collins

LIMS Version: 6.843

Supersedes: JN
Supersedes: JA

Prep Procedure: Po210

Prep Num	LabID	QC Type	Init Alq	Fin Alq	Units	Report Units	Cnt 1 File	Cnt 1 Inst\Det	Cnt 1 Pos Chk By	Cnt 2 File	Cnt 2 Inst\Det	Cnt 2 Pos Chk By	Cnt 3 File	Cnt 3 Inst\Det	Cnt 3 Pos Chk By
1706271-1 PL170616-1PS1							1706271-2 PL170616-1PS2							1706278-2 PL170616-1PS3	
1706278-5 PL170616-1PS4							1706278-6 PL170616-1PS5							1706286-1 PL170616-1PS6	
1706286-3 PL170616-1PS7							1706299-1 PL170616-1PS8							1706299-2 PL170616-1PS9	
1706306-1 PL170616-1PS10							1706306-11 PL170616-1PS11							1706306-12 PL170616-1PS12	
1706328-1 PL170616-1PS13							1706329-2 PL170616-1PS14							1706342-3 PL170616-1PS15	
1706342-3DUP PL170616-1PS16							1706342-7 PL170616-1PS17							1706342-8 PL170616-1PS18	
1706374-1 PL170616-1PS19							1706374-2 PL170616-1PS20							PL170616-1MB PL170616-1PS21	
PL170616-1LCS PL170616-1S22															

Analytical QASS / NCR? Y / N															
Notes															

Reporting Units

LabID:	1stGrpName:	2ndGrpName:	3rdGrpName:	4thGrpName:	5thGrpName:	6thGrpName:	7thGrpName:	8thGrpName:	9thGrpName:	10thGrpName:	11thGrpName:	12thGrpName:	13thGrpName:	14thGrpName:	15thGrpName:	16thGrpName:	17thGrpName:	18thGrpName:	19thGrpName:	20thGrpName:
1706306-1	Po210_USGS																			pCi/l
1706374-1	Po210_USGS																			pCi/l
1706286-1	Po210																			pCi/l
1706299-1	Po210																			pCi/l
1706329-1	Po210																			pCi/l
1706271-1	Po210																			pCi/l
1706271-2	Po210																			pCi/l
1706278-2	Po210_USGS																			pCi/l
1706299-2	Po210																			pCi/l
1706374-2	Po210_USGS																			pCi/l
1706329-2	Po210																			pCi/l
1706286-3	Po210																			pCi/l
1706342-3	Po210_USGS																			pCi/l
1706278-5	Po210_USGS																			pCi/l
1706278-6	Po210_USGS																			pCi/l
1706342-7	Po210_USGS																			pCi/l
1706342-8	Po210_USGS																			pCi/l
1706306-11	Po210_USGS																			pCi/l
1706306-12	Po210_USGS																			pCi/l

Radiochemistry Prep Worksheet

Prep Batch: PL170616-1

Prep Procedure: Po210

Non-Routine Pre-Treatment? Non-Routine
Batch: 04/26/17

Rev: 10

Prep SOP: PAI 711

Prep SOP: NONE

Matrix Class: liquid

Reviewed By: rlm PL Review Date: 6/20/2017
Prep QASS / NCR? 4/21/17Prep Analyst: Tamrae Elhart
Prep Date: 6/16/2017
Prep Dept: AP

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Aliq ml	Fin Aliq ml	Prep Basis	Total Digestate Vol(ml)	Digestate Vol Taken(ml)	Decay Date/Time	Standards	Prep Notes
1	1	1706271-1	SMP	50	25	Filtered	0	0	0	06/19/17 15:50	T1	
2	1	1706271-2	SMP	50	25	Filtered	0	0	0	06/19/17 15:50	T1	
3	1	1706278-2	SMP	187.8788	93.33939	As Received	980	200	0	06/19/17 15:50	T1	
4	1	1706278-5	SMP	188	94	As Received	1000	200	0	06/19/17 15:50	T1	
5	1	1706278-6	SMP	187.8173	93.90863	As Received	985	200	0	06/19/17 15:50	T1	
6	1	1706286-1	SMP	50	25	Filtered	0	0	0	06/19/17 15:50	T1	
7	1	1706286-3	SMP	50	25	Filtered	0	0	0	06/19/17 15:50	T1	
8	1	1706299-1	SMP	50	25	Unfiltered	0	0	0	06/19/17 15:50	T1	
9	1	1706299-2	SMP	50	25	Unfiltered	0	0	0	06/19/17 15:50	T1	
10	1	1706306-1	SMP	187.8788	93.33939	As Received	980	200	0	06/19/17 15:50	T1	
11	1	1706306-11	SMP	187.9397	93.36905	As Received	995	200	0	06/19/17 15:50	T1	
12	1	1706306-12	SMP	188	94	As Received	1000	200	0	06/19/17 15:50	T1	
13	1	1706323-1	SMP	50	25	Unfiltered	0	0	0	06/19/17 15:50	T1	
14	1	1706323-2	SMP	50	25	Unfiltered	0	0	0	06/19/17 15:50	T1	
15	1	1706342-3	SMP	188	94	As Received	1000	200	0	06/19/17 15:50	T1	
16	1	1706342-3	DUP	188	94	As Received	1000	200	0	06/19/17 15:50	T1	
17	1	1706342-7	SMP	186.9565	93.47826	As Received	920	200	0	06/19/17 15:50	T1	
18	1	1706342-8	SMP	187.5	93.75	As Received	960	200	0	06/19/17 15:50	T1	
19	1	1706374-1	SMP	188	94	As Received	1000	200	0	06/19/17 15:50	T1	
20	1	1706374-2	SMP	187.8173	93.90863	As Received	985	200	0	06/19/17 15:50	T1	
21	1	PL170616-1	MB	1000	500	Unfiltered	0	0	0	06/19/17 15:50	T1	
22	1	PL170616-1	LCS	1000	500	Unfiltered	0	0	0	06/19/17 15:50	T1,S1	

Radiochemistry Prep Worksheet

Prep Batch: PL170616-1

Prep Procedure: Po210

Reviewed By: Jim P~

Review Date: 6/20/2017

Non-Routine Pre-Treatment? NBatch: Ses

QA SS

Re-Prep? Y / Batch: Mt

Reviewed By: Jim P~

Review Date: 6/20/2017

Prep SOP: PAI 711 Rev: 10

Batch: Ses

QA SS

Re-Prep? Y / Batch: Mt

Reviewed By: Jim P~

Review Date: 6/20/2017

Prep SOP: NONE

Batch: Ses

QA SS

Re-Prep? Y / Batch: Mt

Reviewed By: Jim P~

Review Date: 6/20/2017

Matrix Class: liquid

Batch: Ses

QA SS

Re-Prep? Y / Batch: Mt

Reviewed By: Jim P~

Review Date: 6/20/2017

Comments _____

Due to potential matrix interferences, reduced aliquots were taken.

Spiked By: Jeremy S. Jones

Date: 6/16/2017

Witnessed By: Lucas A. Dault

Date: 6/16/2017

Tracer/Carrier Solution Information

Soln #	Nuclide	SolnID	Exp Date	Prep Conc	Units	Prep Date	Aliquot	Units	Pipet ID
T1	Po-209	1045.4243.07	22.074	DPM/ml	06/16/17	1	ml	AW026	

Spiked Solution Information									
Soln #	Nuclide	SolnID	Exp Date	Prep Conc	Units	Prep Date	Aliquot	Units	Pipet ID
S1	Pb-210	899.4095.66		94.877	DPM/ml	06/16/17	0.25	ml	AW016
S1	Po-210	899.4095.66		94.877	DPM/ml	06/16/17	0.25	ml	AW016

Prep Procedure: Po210

Prep Batch Not Validated!!!

Reviewed By:

Review Date: _____

Non-Routine Pre-Treatment? Y / N Batch: _____

Re-Prep? Y / N Batch: _____

Prep QASS / NCR? Y / N _____

Prep SOP: PAI 7111 Rev: 10
 Prep SOP: NONE
 Matrix Class: liquid

Prep Analyst: Tambrae Elhart
 Prep Date: 6/16/2017
 Prep Dept: AP

Non-Routine Pre-Treatment? Y / N	Batch: _____	Re-Prep? Y / N	Batch: _____	Prep QASS / NCR? Y / N _____
				Review Date: _____

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Alq ml	Fin Alq ml	Prep Basis	Total Digestate Vol(ml)	Digestate Vol Taken(ml)	Decay Date/Time	Standards	Prep Notes
1	1	1706271-1	SMP	50	50	Unfiltered	0	0	6/19/17	T1		
2	1	1706271-2	SMP	50	50	Unfiltered	0	0	1/5/20	T1		
3	1	1706278-2	SMP	930	187.8788	As Received	990	200		T1		
4	1	1706278-5	SMP	940	188	As Received	1000	200		T1		
5	1	1706278-6	SMP	925	187.8773	As Received	985	200		T1		
6	1	1706286-1	SMP	50	50	Unfiltered	0	0		T1		
7	1	1706286-3	SMP	50	50	Unfiltered	0	0		T1		
8	1	1706295-1	SMP	50	50	Unfiltered	0	0		T1		
9	1	1706295-2	SMP	50	50	Unfiltered	0	0		T1		
10	1	1706306-1	SMP	930	187.8788	As Received	990	200		T1		
11	1	1706306-11	SMP	935	187.9397	As Received	995	200		T1		
12	1	1706306-12	SMP	940	188	As Received	1000	200		T1		
13	1	1706329-1	SMP	50	50	Unfiltered	0	0		T1		
14	1	1706329-2	SMP	50	50	Unfiltered	0	0		T1		
15	1	1706342-3	SMP	940	188	As Received	1000	200		T1		
16	1	1706342-3	DUP	940	188	As Received	1000	200		T1		
17	1	1706342-7	SMP	860	186.9365	As Received	920	200		T1		
18	1	1706342-8	SMP	900	187.5	As Received	960	200		T1		
19	1	1706374-1	SMP	940	188	As Received	1000	200		T1		
20	1	1706374-2	SMP	925	187.8773	As Received	985	200		T1		
21	1	PL170616-1	MB	1000	1000	Unfiltered	0	0		T1		
22	1	PL170616-1	LCS	1000	1000	Unfiltered	0	0		T1,S1		

Prep Procedure: Po210

Prep Batch Not Validated!!!

Reviewed By:

Review Date:

Non-Routine Pre-Treatment? Y / N	Batch: _____	Re-Prep? Y / N	Batch: _____	Prep QASS / NCR? Y / N
Prep SOP: PAI 711 Rev: 10	Prep Analyst: Tambrae Elhart			Balance:
Prep SOP: NONE	Prep Date: 6/16/2017			Balance:
Matrix Class: liquid	Prep Dept: AP			

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Alq ml	Prep Basis	Total Digestate Vol/ml	Digestate Vol Taken/ml	Decay Date/Time	Standards	Prep Notes

Comments

Due to potential matrix interferences, reduced aliquots were taken.

Spiked By: TST Date: 6/16/17Witnessed By: JK Doss Date: 6/16/17

Tracer/Carrier Solution Information

Soln #	Nuclide	SolnID	Exp Date	Prep Conc	Units	Prep Date	Aliquot	Units	Prep Date	Aliquot	Units	Pipet ID
T1	Po-209	10454243.07	8/3/17	22.074	DFM/ml	06/16/17	1	ml	06/16/17	94.877	ml	AW026
										94.877		

Spike Solution Information												
Soln #	Nuclide	SolnID	Exp Date	Prep Conc	Units	Prep Date	Aliquot	Units	Prep Date	Aliquot	Units	Pipet ID
S1	Pb-210	8994095.66	10/21/17	94.877	DFM/ml	06/16/17	0.25	ml	06/16/17	94.877	ml	AW016
S1	Po-210	8994095.66										AW016

Sample Condition Form (Liquid)

Analyst: JSS

Analysis Date: 6/16/17

Method: Prep

Work Order	Sample ID	Sample Condition (Visual Appearance of Analysis Aliquot at Time of Prep)			Remarks
		pH	Color		
1706271	1	<2	Tan	Sedimented (Filtered, Filtered)	
↓	2	≈7	↓	↓	+ (Acidified)
1706278	2	<2	Clear	None	
↓	5		↓	↓	
↓	6		↓	↓	
1706286	1		Tan	Sedimented (Filtered, Filtered)	
↓	3		Yellow	↓	
1706299	1		Clear	Dusting of sediment	
↓	2			↓	
1706306	1			None	
↓	11			↓	
↓	12	↓		↓	
1706329	1	≈7		Dusting of sediment (Acidified)	
↓	2	≈3		↓	
1706342	3	<2		None	
↓	7	↓		↓	
↓	8	↓		↓	
1706374	1				
↓	2	↓	↓	↓	

Batch: PL170616-1

Po Solid

Reagent	Lot #
1N Hydrochloric Acid	0000132881
Ascorbic Acid	0000120890
Amyl Acetate	K46604
Flexible Colloidian	0000087213

Po Liquid

Reagent	Lot #
Conc. Hydrochloric Acid	0000132881
1N Hydrochloric Acid	0000132881
Ascorbic Acid	0000120890
Amyl Acetate	K46604
Flexible Colloidian	0000087213

Reporting Units

Unit	Instrument Name	RptUnits:
1706426-1	Po210	pCi/l
1706423-1	Po210	pCi/l
1706421-1	Po210_USGS	pCi/l
1706341-1	Po210	pCi/l
1706426-3	Po210	pCi/l
1706341-3	Po210	pCi/l
1706286-3	Po210	pCi/l

Radiochemistry Prep Worksheet

Prep Batch: PL170623-1

Prep Procedure: Po210

Non-Routine Pre-Treatment? Y / N

Batch: A1A

Re-Prep? Y / N

Batch: N/A

Prep SOP: PAI 711 Rev: 10

Prep SOP: NONE

Matrix Class: liquid

Reviewed By: lad LAD

Review Date: 6/25/2017

Prep Analyst: Lucas A. Daut LAD									
Prep Date: 6/23/2017									
Prep Dept: AP									

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Alq ml	Fin Alq ml	Prep Basis	Total Digestate Vol(ml)	Digestate Vol Taken(ml)	Decay Date/Time	Standards	Prep Notes
1	1	1706286-3	SMP	10	5	Unfiltered	0	0	0	06/24/17 16:08	T1	
2	1	1706341-1	SMP	50	25	Unfiltered	0	0	0	06/24/17 16:08	T1	
3	1	1706341-3	SMP	50	25	Unfiltered	0	0	0	06/24/17 16:08	T1	
4	1	1706421-1	SMP	187.8788	93.9394	As Received	990	200	0	06/24/17 16:08	T1	
5	1	1706421-1	DUP	187.8788	93.9394	As Received	990	200	0	06/24/17 16:08	T1	
6	1	1706423-1	SMP	50	25	Unfiltered	0	0	0	06/24/17 16:08	T1	
7	1	1706423-1	SMP	50	25	Unfiltered	0	0	0	06/24/17 16:08	T1	
8	1	1706426-3	SMP	50	25	Unfiltered	0	0	0	06/24/17 16:08	T1	
9	1	PL170623-1	MB	1000	500	Unfiltered	0	0	0	06/24/17 16:08	T1	
10	1	PL170623-1	LCS	1000	500	Unfiltered	0	0	0	06/24/17 16:08	T1,S1	

Comments

Due to potential matrix interference reduced aliquots were taken; MDCs were met.

Spiked By: Lucas A. Daut Date: 6/23/2017

Witnessed By: Tamrae Elhart Date: 6/23/2017

Tracer/Carrier Solution Information

Soln #	Nuclide	SolnID	Exp Date	Prep Conc.	Units	Prep Date	Aliquot Units	Prep Date	Aliquot Units	Pipet ID
T1	Po-209	1045.4243.07	8/3/17	22.072	DPM/ml	06/23/17	1 ml	06/23/17	0.25 ml	Aw016

Spike Solution Information										
Soln #	Nuclide	SolnID	Exp Date	Prep Conc.	Units	Prep Date	Aliquot Units	Prep Date	Aliquot Units	Pipet ID
S1	Pb-210	899.4095.66	10/21/17	94.822	DPM/ml	06/23/17	0.25 ml	06/23/17	0.25 ml	Aw016
S1	Po-210	899.4095.66		94.822	DPM/ml	06/23/17	0.25 ml	06/23/17	0.25 ml	Aw016

Prep Procedure: Po210

Prep Batch Not Validated!!

Reviewed By:

Non-Routine Pre-Treatment? Y / N Batch: _____
 Prep SOP: PAI 711 Rev: 10
 Prep SOP: NONE
 Matrix Class: liquid

Re-Prep? Y / N Batch: _____
 Prep QASS / NCR? Y / N _____

Review Date:

Prep Analyst: Lucas A. Daut
 Prep Date: 6/23/2017
 Prep Dept: AP

Balance:
 Balance:

Samp Num	Prep Num	LabID	QC Type	Dish No	Init Alq ml	Fin Alq ml	Prep Basis	Total Digestate Vol(ml)	Digestate Vol Taken(ml)	Decay Date/Time	Standards	Prep Notes
1	1	1706286-3	SMP	50	210	1000	1000	0	0	4:08 PM	T1	Matrix Interfere
2	1	1706341-1	SMP	50	210	1000	1000	0	0	6/24	T1	
3	1	1706341-3	SMP	210	210	1000	1000	0	0		T1	MDC's are met
4	1	1706421-1	SMP	910	187.8788	As Received	990	200			T1	
5	1	1706421-1	DUP	930	187.8788	As Received	990	200			T1	
6	1	1706423-1	SMP	210	210	1000	1000	0	0		T1	Matrix Interfere
7	1	1706426-1	SMP	210	210	1000	1000	0	0		T1	
8	1	1706426-3	SMP	210	210	1000	1000	0	0		T1	
9	1	PL170623-1	MB	1000	1000	1000	1000	0	0		T1	
10	1	PL170623-1	LCS	1000	1000	1000	1000	0	0		T1,51	

Comments

[Signature] Spiked By: Lucas A. Date: 6/23/17
 Witnessed By: John A. Date: 6/23/17

Reviewed By:

Spike Solution Information												
Soln #	Nuclide	SolnID	Exp Date	Prep Conc	Units	Prep Date	Aliquot	Units	Prep Date	Aliquot	Units	Pipet ID
T1	Po-209	1045-4243-07	8/3/17	22.072	DPM/ml	06/23/17	1	ml	AW026	1	94.822	DFN/ml
											94.822	DFN/ml

Tracer/Carrier Solution Information												
Soln #	Nuclide	SolnID	Exp Date	Prep Conc	Units	Prep Date	Aliquot	Units	Prep Date	Aliquot	Units	Pipet ID
T1	Po-209	1045-4243-07	8/3/17	22.072	DPM/ml	06/23/17	1	ml	AW026	1	94.822	DFN/ml

PL170623-1 LAD
6/25

Batch: PL170616-1

Po Solid

Reagent	Lot #
1N Hydrochloric Acid	0000132881
Ascorbic Acid	0000120890
Amyl Acetate	K46604
Flexible Colloidian	0000087213

Po Liquid

Reagent	Lot #
Conc. Hydrochloric Acid	0000132881
1N Hydrochloric Acid	0000132881
Ascorbic Acid	0000120890
Amyl Acetate	K46604
Flexible Colloidian	0000087213

Section 8

STANDARDS TRACEABILITY DOCUMENTS

8

Pb-210 899.4095.66 working standard

1E 12/8/14

Prepare a working dilution of 899.3610.421. Density of 1M HNO₃, lot # 0000084136Mass of 100mL vol. flask: 56.4468gBalance # 12Mass of flask & 100mL acid: 159.4521gBalance# 12Net Mass: 103.0053gDensity: 1.0301g/ml2. Mass of 899.3610.42 transferred:Mass of open empty nalgene: 74.5139gBalance# 12Mass of nalgene & standard: 77.1985gBalance# 12Net mass of standard transferred: 2.6846gBalance# N/A18
12/8/14

3. Dilute to final volume:

Mass of nalgene, standard, & diluent: 1147.7g Balance# 26Mass of empty nalgene (from above): 74.5139g Balance# 12Net mass of new dilution: 1073.1861g Balance# N/A

4. Final activity calculation:

$$\frac{46,996.4 \text{ dpm/g} (1.0301\text{g/ml})(2.6846\text{g})}{(1073.1861\text{g})} = 121.10 \text{ dpm/ml}$$

18
12/8/14

JP 11/3/15

Stnd ID: 899.4095.66

Description: Pb-210

Expiration: 10/26/2016

Activity: 121.10 dpm/ml

2s Uncertainty: 3.39 dpm/ml

Ref. Date: 8/10/2009

Ref Time: N/A

Prep Date: 12/8/2014 Prep by: TE

Matrix/Comp. 1 M HNO₃

Half Life (y): 2.22E+01

Reverification Log		
Analysis Date	Initials	Expiration Date
10/21/2016	JP	10/21/2017

JP 11/3/15

Continued on Page

1 Elliott

Signed

12/8/14

Date

Read and Understood By
J. Elliott

Signed

11/3/15

Date 48 of 71



Eckert & Ziegler

Analytics

PSO#
899 Rel 6/14/09

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318
Tel 404-352-8677
Fax 404-352-2837
www.analyticsinc.com

CERTIFICATE OF CALIBRATION

Standard Radionuclide Source

80328-307

Pb-210 5 mL Liquid in Flame Sealed Vial

Customer: ALS Laboratory Group / Fort Collins
P.O. No.: 73625 07-24-09, Item 1

This standard radionuclide source was prepared gravimetrically from a calibrated master solution. The master solution was calibrated by liquid scintillation counting. Radionuclide purity and calibration were checked by germanium gamma-ray spectrometry and liquid scintillation counting. The nuclear decay rate and assay date for this source are given below. Eckert & Ziegler Analytics (EZA) maintains traceability to the National Institute of Standards and Technology through a Measurements Assurance Program as described in USNRC Regulatory Guide 4.15, Revision 1, February, 1979, and compliance with ANSI N42.22-1995, "Traceability of Radioactive Sources to NIST." EZA is accredited by the Health Physics Society (HPS) for the production of NIST-traceable sources, and this source was produced in accordance with the HPS accreditation requirements. Customers may report any concerns with the accreditation program to the HPS Secretariat, 1313 Dolley Madison Blvd., Ste. 402, McLean, VA 22101.

Isotope	Half-Life, Days	Activity (Bq)	Uncertainty*, %			Reference Date (12:00 PM EST)
			Type u_A	Type u_B	Type U	
Pb-210	8145.1	4.078E+04	0.1	1.4	2.8	08/10/2009

*Uncertainty: U - Relative expanded uncertainty, $k = 2$. See NIST Technical Note 1297, "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results."

Comments:

Impurities: gamma-impurities <0.1%, alpha-impurities (other than decay products) <0.01%. 52.06357 grams 1M HNO₃ solution, carrier free.


Source Prepared by:

N. E. Kasate, Radiochemist

QA Approved:

D. M. Montgomery, QA Manager

Date: 8-13-09



Corporate Office

24937 Avenue Tibbitts Valencia, California 91355

Laboratory

1380 Seaboard Industrial Blvd. Atlanta, Georgia, 30314 **of 71**

Prepare a working dilution of RSO# 1045

18 7/31/16

1. Density of 2M HCl, lot # 0606128723Mass of 100mL vol. flask: 56.4426gMass of flask & 100mL acid: 159.7513gNet Mass: 103.3087gDensity: 1.03309g/mlBalance # 12Balance# 122. Mass of RSO# 1045 transferred:Mass of open empty nalgene: 74.8232gMass of nalgene & standard: 79.8830gNet mass of standard transferred: 5.0598gBalance# 12Balance# 1218
7/31/16

3. Dilute to final volume:

Mass of nalgene, standard, & diluent: 1171.2gMass of empty nalgene (from above): 74.8232gNet mass of new dilution: 1096.3768gBalance# 26Balance# 12

4. Final activity calculation:

$$400.3 \text{ Bq} \frac{(1.03309 \text{ g/mL}) (5.0598 \text{ g}) (60 \text{ dpm})}{(1096.3768 \text{ g}) (5.1515 \text{ g}) (1 \text{ Bq})} = 22.23 \text{ dpm/mL}$$

18
7/31/16Stnd ID: **1045.4243.07**Description: Po-209Expiration: 8/3/2017Activity: 22.23 dpm/mL2s Uncertainty: 0.44 dpm/mLRef. Date: 6/8/2016Ref Time: N/APrep Date: 7/31/2016 Prep by: TEMatrix/Comp. 2.0 M HClHalf Life (y): 1.02E+02

Reverification Log		
Analysis Date	Initials	Expiration Date

Continued on Page _____

J. Elliott

Signed

7/31/16

Date

Read and Understood By_____
Signed08/08/16

Date

50 of 71



Eckert & Ziegler

Analytics

R50 #
1045

Rec'd
6-10-16

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318
Tel 404-352-8677
Fax 404-352-2837
www.ezag.com

CERTIFICATE OF CALIBRATION

Standard Reference Source

SRS Number: 103361

Source Description: 5 mL Liquid in Flame Sealed Vial

Product Code: 8209

Customer: ALS Laboratory Group

P.O. Number: FC001053, Item 1

This standard radionuclide source was prepared gravimetrically from a master solution calibrated by Eckert & Ziegler Analytics (EZA). The master solution was calibrated by liquid scintillation counting. Radionuclide calibration and purity were checked by germanium gamma-ray spectrometry, liquid scintillation counting, and/or alpha spectrometry, as applicable. The nuclear decay rate and reference date for this source are given below. EZA maintains traceability to the National Institute of Standards and Technology (NIST) through a Measurements Assurance Program as described in USNRC Regulatory Guide 4.15, Revision 2, July 2007, and compliance with ANSI N42.22-1995, "Traceability of Radioactive Sources to NIST."

Reference Date: 08-June-2016 12:00 PM EST

Isotope	Half-Life, d	Activity, Bq	Uncertainty			Calibration Method**
			u_A , %	u_B , %	U , %*	
Po-209	3.726E+04	4.003E+02	0.3	1.0	2.0	4π LS

Uncertainty:** U - Relative expanded uncertainty, $k = 2$. See NIST Technical Note 1297, "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results." *Calibration Methods:** 4π LS - 4π Liquid Scintillation Counting, HPGe - High Purity Germanium Gamma-Ray Spectrometer, IC - Ionization Chamber.

(Certificate continued on reverse side)

SRS Number: 103361

Comments:


5.15159 g 2 M HCl solution, carrier free

Impurities:

γ -impurities < 0.1%, α -impurities < 0.1%


This source was wipe tested in its inactive areas with leak test results < 185 Bq (5 nCi) of removable activity per ISO 9978:1992.

Source Prepared by:



K. Hardley, Radiochemist

QC Approved by:



J. Lahr, Spectroscopist

Date: 02-JUN-16

Section 9

ADDITIONAL SUPPORTING DOCUMENTATION

Alpha Spectroscopy

Quality Control Data

Weekly Background, Energy, and Efficiency Calibrations

Calibration Data Summary

Laboratory Name: ALS -- Fort Collins
 PAI Work Order: 1706286

Prep SOP: PAI 711
 Analytical SOP: PAI 714

Reported on: Monday, July 03, 2017
 2:29:56 PM

Lab Sample ID Spectrum Analysis Date	QC Type	Batch ID Analysis Run	Test Name	Detector Id	Eff Spectrum Bkg Spectrum Egy Spectrum	Eff Date Bkg Date Egy Date	RESULTS %Efficiency Bkg CPM Energy keV	FLAGS Efficiency Background Energy	LCL %Efficiency Bkg CPM Energy keV	UWL %Efficiency Bkg CPM Energy keV	UCL %Efficiency Bkg CPM Energy keV
1706286-1 6/21/2017	SMP	PL170616-1 PL170616-1A	Po210	105	C170620105A B170620105	6/20/2017 6/20/2017	31.73 0.0260	Pass Pass	31.09 0.0000	31.65 0.0100	33.81 0.1000
					C170620105A	6/20/2017	5555.8	Pass	5506.0	5516.0	5596.0
1706286-3 6/25/2017	SMP	PL170623-1 PL170623-1A	Po210	103	C170620103 B170620103	6/20/2017 6/20/2017	30.78 0.0180	Pass Pass	30.23 0.0000	30.77 0.0100	32.87 0.1000
					C170620103	6/20/2017	5555.8	Pass	5506.0	5516.0	5596.0
PL170616-1 6/21/2017	MB	PL170616-1 PL170616-1A	Po210	113	C170620113 B170620113	6/20/2017 6/20/2017	32.54 0.0180	Warning Pass	32.00 0.0000	32.58 0.0100	34.80 0.1000
					C170620113	6/20/2017	5555.8	Pass	5506.0	5516.0	5596.0
PL170616-1 6/21/2017	LCS	PL170616-1 PL170616-1A	Po210	114	C170620114 B170620114	6/20/2017 6/20/2017	33.42 0.0110	Pass Pass	31.25 0.0000	32.80 0.0100	33.89 0.1000
					C170620114	6/20/2017	5555.8	Pass	5506.0	5516.0	5596.0
PL170623-1 6/25/2017	MB	PL170623-1 PL170623-1A	Po210	112	C170620112 B170620112	6/20/2017 6/20/2017	31.90 0.0140	Warning Pass	31.79 0.0000	32.36 0.0100	33.89 0.1000
					C170620112	6/20/2017	5555.8	Pass	5506.0	5516.0	5596.0
PL170623-1 6/25/2017	LCS	PL170623-1 PL170623-1A	Po210	113	C170620113 B170620113	6/20/2017 6/20/2017	32.54 0.0180	Warning Pass	32.00 0.0000	32.58 0.0100	34.80 0.1000
					C170620113	6/20/2017	5555.8	Pass	5506.0	5516.0	5596.0

Data Package ID: PL1706286-1

Abbreviations:	Eff - Efficiency Egy - Energy	Bkg - Background CPM - Counts per Minute	LCL - Lower Control Limit LWL - Lower Warning Limit	UWL - Upper Warning Limit
5			UCL - Upper Control Limit	CI - The Analysis Date exceeds the Calibration Date by more than 14 days.

Date Printed: Wednesday, July 05, 2017

ALS -- Fort Collins

LIMS Version: 6.843

Page 1 of 1

Alpha Spec Calibration Source Re-Certification

Recalibration performed by Isotope Products Laboratories

Primary Certified Source

Source PA ID:	190
Planchet Label:	9
Recalibrated on:	10/4/2016
Received by AIS on:	10/16/2013

Values from certificate						
Source ID:	92MIX223027					
Total Activity:	3745.2	dpm				
Ref. Date:	10/15/2013					
Nuclide	Act (Bq)	Act (dpm)	Half-life (yrs)	Decay Corrected		
U-234	49.54	2972.4	2.48E+05	2972.38 dpm		
U-235	1.09	65.58	7.0E+08	65.58 dpm		
Am-241	11.79	707.4	432.17	704.04 dpm		
		TOTAL		3741.98 dpm		

Efficiency Determination for Detector:

Source Serial# PA ID Sequential # Count Date Am-241 U-234 U-235 Count dur Known Detector efficiency

Source Serial#	PA ID	Sequential #	Count Date	Am-241	U-234	U-235	Count	dur	Known dpm	Detector efficiency
92MIX223027	190	97-19-103-09	10/4/16	7562	32112	1070	2100	1162.40	3741.98	31.06%

Sources 1 through 8 activity determination

Source Serial#	PA ID	Sequential #	Count Date	Am-241	U-234	U-235	Count	dur	Known dpm	Detector efficiency	Combined dpm
92MIX223026	182	97-19-103-01	10/4/16	79881	21777	2100	31.06%	1193.95	7343.17	255.42	8792.54
92MIX223028	183	97-19-103-02	10/4/16	15085	148128	3063	2100	31.06%	1387.47	13824.37	1567.15
92MIX223024	184	97-19-103-03	10/4/16	67474	70483	2808	2100	31.06%	6206.06	6482.82	238.88
92MIX223021	185	97-19-103-04	10/4/16	21981	60440	2557	2100	31.06%	2019.91	5558.09	235.19
92MIX223025	186	97-19-103-05	10/4/16	97883	114458	3780	2100	31.06%	9002.99	10527.51	784.16
92MIX223022	187	97-19-103-06	10/4/16	72777	79883	2564	2100	31.06%	747.40	237.67	1978.16
92MIX223023	188	97-19-103-07	10/4/16	43617	68053	2043	2100	31.06%	4011.76	6342.09	187.91
92MIX223029	189	97-19-103-08	10/4/16	33986	214074	7195	2100	31.06%	3124.89	19689.89	661.77

Efficiency Verification

Source Serial#	PA ID	Sequential #	Count Date	Am-241	U-234	U-235	Count	dur	Known dpm	Detector efficiency	RPD	FLAG
92MIX223027	190	97-19-103-09	10/5/16	7807	32611	1278	2100	1191.31	3741.99	31.84%	-2.46%	PASS

Sources 1 through 8 activity re-verification

Source Serial#	PA ID	Sequential #	Count Date	Am-241	U-234	U-235	Count	dur	Known dpm	Detector efficiency	Combined dpm	Within % of Certified value?
92MIX223026	182	97-19-103-01	8792.54	8849.86	-0.65%	Yes						
92MIX223028	183	97-19-103-02	15367.15	15692.36	-3.91%	Yes						
92MIX223024	184	97-19-103-03	12328.75	13503.77	-4.26%	Yes						
92MIX223021	185	97-19-103-04	7814.18	8161.24	-4.25%	No						
92MIX223025	186	97-19-103-05	19878.16	20876.95	-5.25%	No						
92MIX223022	187	97-19-103-06	14278.88	15285.63	-6.59%	Yes						
92MIX223023	188	97-19-103-07	10541.76	10723.95	-1.70%	Yes						
92MIX223029	189	97-19-103-08	23475.75	23593.84	-0.50%	Yes						

* Certificate values decay corrected to the count date

Ok 10/5/16

10/4/2017

Data from certificates

Reference Date	U-234 (Bq)	U-234 (dpm)	U-235 (Bq)	U-235 (dpm)	Am-241 (Bq)	Am-241 (dpm)
5/1/2003	124.10	7446.00	2.43	145.74	21.43	1285.80
5/1/2003	238.30	14356.00	4.20	252.00	23.35	1413.00
5/1/2003	119.40	7184.00	1.93	115.56	106.00	6380.00
4/1/2003	101.00	6080.00	1.26	75.84	34.50	2070.00
4/1/2003	203.00	12180.00	3.41	146.40	8784.00	
4/1/2003	132.90	7974.00	3.17	108.96	121.30	7278.00
4/1/2003	107.10	6426.00	0.93	55.54	72.26	4335.60
5/1/2003	334.80	20088.00	6.55	393.18	53.02	3181.20

Analyst: ORTEC

Detector: 129

9:14:40AM 10/5/2016

Energy Calibration: SOURCE190_10.04.16 (#9)

CalibrationAnalysis Date: 10/4/2016 12:09:56PM
Calibration Type: Energy And Efficiency

Description:

Source Info

Certification Date: 10/15/2013 10:44:40AM

Certificate ID: A9 RSO#190

Prepared by: Isotope Product Laboratories

Description:

Acquisition

Detector: 129, SN:5505430, ID: 129

Energy Calibration Equation:

Acquisition Start Date: 10/4/2016 11:26:06AM

Gain = 9.9003 keV / Ch

Live Time: 35.00 min.


Offset = 3,021.28 keV

Real Time: 35.01 min.

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE190_10.04.16 (#9)

Efficiency: 33.86% +/- 1.39% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Initial Calibration: Yes

Algorithm: Linear

Shelf: 0

Nuclide Activity Summary

Nuclide	Peak Channel	Peak Energy keV	ROI Start Channel	ROI End Channel	Peak FWHM keV	Gross Counts	Net Count Rate (cpm)
U-235	139	4,396.00	100	152	0.00	1,070.00	33.71
U-234	177	4,775.80	153	190	68.12	32,112.00	953.94
Am-241	249	5,485.70	210	270	71.83	7,502.00	221.80

JP 10/4/16

Analyst: ORTEC

Detector: 129

12:49:08PM 10/4/2016

Energy Calibration: SOURCE182_10.04.16 (#1)

CalibrationAnalysis Date: 10/4/2016 12:47:23PM
Calibration Type: Energy And Efficiency

Description:

Source Info

Certificate ID: A1 RSO#182

Certification Date: 5/1/2003 10:27:02AM

Prepared by: Isotope Product Laboratories

Description:

Acquisition

Detector: 129, SN:5505430, ID: 129

Energy Calibration Equation:

Acquisition Start Date: 10/4/2016 12:11:42PM

Gain = 9.9003 keV / Ch

Live Time: 35.00 min.


Offset = 3,021.28 keV

Real Time: 35.02 min.

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE182_10.04.16 (#1)

Efficiency: 33.04% +/- 2.03% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Initial Calibration: Yes

Algorithm: Linear

Shelf: 0

Nuclide Activity Summary

Nuclide	Peak Channel	Peak Energy keV	ROI Start Channel	ROI End Channel	Peak FWHM keV	Gross Counts	Net Count Rate (cpm)
U-235	139	4,396.00	100	152	84.43	2,777.00	79.34
U-234	177	4,775.80	153	190	71.29	79,837.00	2,281.06
Am-241	249	5,485.70	210	270	72.86	12,981.00	370.89

JP 10/4/16

Analyst: ORTEC

Detector: 129

1:25:35PM 10/4/2016

Energy Calibration: SOURCE183_10.04.16 (#2)

CalibrationAnalysis Date: 10/4/2016 1:25:29PM
Calibration Type: Energy And Efficiency

Description:

Source Info

Certificate ID: A2 RSO#183

Certification Date: 5/1/2003 10:33:40AM

Prepared by: Isotope Product Laboratories

Description:

Acquisition

Detector: 129, SN:5505430, ID: 129

Energy Calibration Equation:

Acquisition Start Date: 10/4/2016 12:49:04PM

Gain = 9.9003 keV / Ch

Live Time: 35.00 min.


Offset = 3,021.28 keV

Real Time: 35.03 min.

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE183_10.04.16 (#2)

Efficiency: 32.74% +/- 1.25% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Initial Calibration: Yes

Algorithm: Linear

Shelf: 0

Nuclide Activity Summary

Nuclide	Peak Channel	Peak Energy keV	ROI Start Channel	ROI End Channel	Peak FWHM keV	Gross Counts	Net Count Rate (cpm)
U-235	139	4,396.00	100	152	75.93	3,863.00	110.37
U-234	177	4,775.80	153	190	69.70	148,128.00	4,232.23
Am-241	249	5,485.70	210	270	73.20	15,085.00	431.00

JP 10/4/16

Analyst: ORTEC

Detector: 129

2:17:25PM 10/4/2016

Energy Calibration: SOURCE184_10.04.16 (#3)

CalibrationAnalysis Date: 10/4/2016 2:16:56PM
Calibration Type: Energy And Efficiency

Description:

Source Info

Certificate ID: A3 RSO#184

Certification Date: 5/1/2003 10:36:52AM

Prepared by: Isotope Product Laboratories

Description:

Acquisition

Detector: 129, SN:5505430, ID: 129

Energy Calibration Equation:

Acquisition Start Date: 10/4/2016 1:26:53PM

Gain = 9.9003 keV / Ch

Live Time: 35.00 min.


Offset = 3,021.28 keV

Real Time: 35.03 min.

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE184_10.04.16 (#3)

Efficiency: 31.83% +/- 1.26% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Initial Calibration: Yes

Algorithm: Linear

Shelf: 0

Nuclide Activity Summary

Nuclide	Peak Channel	Peak Energy keV	ROI Start Channel	ROI End Channel	Peak FWHM keV	Gross Counts	Net Count Rate (cpm)
U-235	139	4,396.00	100	152	0.00	2,608.00	74.51
U-234	177	4,775.80	153	190	74.20	70,483.00	2,013.80
Am-241	249	5,485.70	210	270	74.76	67,474.00	1,927.83

JP 10/4/16

Analyst: ORTEC

Detector: 129

9:00:58AM 10/5/2016

Energy Calibration: SOURCE185_10.04.16 (#4)

CalibrationAnalysis Date: 10/4/2016 2:53:33PM
Calibration Type: Energy And Efficiency

Description:

Source Info

Certificate ID: A4 RSO#185

Certification Date: 4/1/2003 10:38:09AM

Prepared by: Isotope Product Laboratories

Description:

Acquisition

Detector: 129, SN:5505430, ID: 129

Energy Calibration Equation:

Acquisition Start Date: 10/4/2016 2:18:14PM

Gain = 9.9003 keV / Ch

Live Time: 35.00 min.


Offset = 3,021.28 keV

Real Time: 35.02 min.

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE185_10.04.16 (#4)

Efficiency: 31.73% +/- 1.30% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Initial Calibration: Yes

Algorithm: Linear

Shelf: 0

Nuclide Activity Summary

Nuclide	Peak Channel	Peak Energy keV	ROI Start Channel	ROI End Channel	Peak FWHM keV	Gross Counts	Net Count Rate (cpm)
U-235	139	4,396.00	100	152	79.20	2,557.00	73.06
U-234	177	4,775.80	153	190	80.97	60,440.00	1,726.86
Am-241	249	5,485.70	210	270	77.99	21,961.00	627.46

30 10/4/16

Analyst: ORTEC

Detector: 129

7:04:12AM 10/5/2016

Energy Calibration: SOURCE188_10.04.16 (#7)

CalibrationAnalysis Date: 10/5/2016 7:03:07AM
Calibration Type: Energy And Efficiency

Description:

Source Info

Certificate ID: A7 RSO#188

Certification Date: 4/1/2003 10:42:01AM

Prepared by: Isotope Product Laboratories

Description:

Acquisition

Detector: 129, SN:5505430, ID: 129

Energy Calibration Equation:

Acquisition Start Date: 10/5/2016 6:13:44AM

Gain = 9.9003 keV / Ch

Live Time: 35.00 min.


Offset = 3,021.28 keV

Real Time: 35.02 min.

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE188_10.04.16 (#7)

Efficiency: 31.88% +/- 1.31% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Initial Calibration: Yes

Algorithm: Linear

Shelf: 0

Nuclide Activity Summary

Nuclide	Peak Channel	Peak Energy keV	ROI Start Channel	ROI End Channel	Peak FWHM keV	Gross Counts	Net Count Rate (cpm)
U-235	139	4,396.00	100	152	77.73	2,043.00	58.37
U-234	177	4,775.80	153	190	69.01	68,953.00	1,970.09
Am-241	249	5,485.70	210	270	71.83	43,617.00	1,246.20

10/4/16

Analyst: ORTEC

Detector: 129

7:44:05AM 10/5/2016

Energy Calibration: SOURCE189_10.04.16 (#8)

CalibrationAnalysis Date: 10/5/2016 7:43:56AM
Calibration Type: Energy And Efficiency

Description:

Source Info

Certificate ID: A8 RSO#189

Certification Date: 5/1/2003 10:43:18AM

Prepared by: Isotope Product Laboratories

Description:

Acquisition

Detector: 129, SN:5505430, ID: 129

Energy Calibration Equation:

Acquisition Start Date: 10/5/2016 7:04:08AM

Gain = 9.9003 keV / Ch

Live Time: 35.00 min.


Offset = 3,021.28 keV

Real Time: 35.05 min.

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE189_10.04.16 (#8)

Efficiency: 33.82% +/- 1.28% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Initial Calibration: Yes

Algorithm: Linear

Shelf: 0

Nuclide Activity Summary

Nuclide	Peak Channel	Peak Energy keV	ROI Start Channel	ROI End Channel	Peak FWHM keV	Gross Counts	Net Count Rate (cpm)
U-235	139	4,396.00	100	152	87.84	7,195.00	205.57
U-234	177	4,775.80	153	190	74.72	214,074.00	6,116.40
Am-241	249	5,485.70	210	270	74.87	33,966.00	970.46

10/4/16

Analyst: ORTEC

Detector: 129

8:36:03AM 10/5/2016

CalibrationEnergy Calibration: SOURCE190A_10.04.16 (#9)
Description:Analysis Date: 10/5/2016 8:35:09AM
Calibration Type: Energy And Efficiency**Source Info**Certificate ID: A9 RSO#190
Prepared by: Isotope Product Laboratories
Description:

Certification Date: 10/15/2013 10:44:40AM


AcquisitionDetector: 129, SN:5505430, ID: 129
Acquisition Start Date: 10/5/2016 7:45:09AMEnergy Calibration Equation:
Gain = 9.9003 keV / ChLive Time: 35.00 min.
Real Time: 35.01 min.

Offset = 3,021.28 keV

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE190A_10.04.16 (#9)

Efficiency: 33.67% +/- 1.38% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Initial Calibration: Yes

Algorithm: Linear

Shelf: 0

Nuclide Activity Summary

Nuclide	Peak Channel	Peak Energy keV	ROI Start Channel	ROI End Channel	Peak FWHM keV	Gross Counts	Net Count Rate (cpm)
U-235	139	4,396.00	100	152	0.00	1,278.00	36.51
U-234	177	4,775.80	153	190	68.47	32,611.00	931.74
Am-241	249	5,485.70	210	270	71.17	7,807.00	223.06

10/5/16
JW



Eckert & Ziegler

Isotope Products

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661•309•1010
Fax 661•257•8303

#190
Received 10/18/13

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

Radionuclide:	U-234	Customer:	ALS LABORATORY
Radionuclide:	U-235	P.O. No.:	FC 3595 / R5576
Radionuclide:	Am-241	Catalog No.:	*SOURCE-RECAL-STD
Half-life (U-234):	(2.454 ± 0.006)E+05 years	Reference Date:	15-Oct-13 12:00 PST
Half-life (U-235):	(7.037 ± 0.011)E+08 years	Source No.:	92MIX223027
Half-life (Am-241):	432.17 ± 0.66 years		

Contained Radioactivity:

U-234:	1.339 nCi,	49.54 Bq	Am-241:	0.3187 nCi,	11.79 Bq
U-235:	0.02954 nCi,	1.093 Bq	Total Activity:	1.687 nCi,	62.42 Bq

Physical Description:

- A. Capsule type: Disk (22 mm OD x 0.79 mm THK)
- B. Nature of active deposit: Electrodeposited and diffusion bonded oxide
- C. Active diameter/volume: 19 mm
- D. Backing: Stainless steel
- E. Cover: None

Radioimpurities: Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in May 2001.

Uncertainty of Measurement:

- A. Type A (random) uncertainty: ± 0.5 %
- B. Type B (systematic) uncertainty: ± 3.0 %
- C. Uncertainty in aliquot weighing: ± 0.0 %
- D. Total uncertainty at the 99% confidence level: ± 3.0 %

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (as in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 1893 α/min in 2π on 20-Sep-13.

Daniel James Van Dalsen
Quality Control

2-OCT-13
Date

IPL Ref. No.: 987-28

ISO 9001 CERTIFIED

Medical Imaging Laboratory
24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street Burbank, California 91504



An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661•309•1010
Fax 661•257•8303

Re Calibrated 10/4/16
New Exp Date 10/4/2017
PAI 187
recalibrated 4-15-03
TM10516

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW040203/R2193
Catalog No.: MISC-STD
Reference Date: 1-May-03 12:00 PST
Source No.: 92MIX2203026

Contained Radioactivity:

U-234:	3.354 nCi (124.1 Bq)	Am-241:	0.5793 nCi (21.43 Bq)
U-235:	0.06566 nCi (2.429 Bq)	Total Activity:	3.999 nCi (148.0 Bq)

Physical description:

A. Capsule type: Disk (22 mm OD X 0.79 mm THK)
B. Nature of active deposit: Electrodeposited and diffusion bonded oxides
C. Active Diameter: 19 mm
D. Backing: Stainless steel
E. Cover: None

Radioimpurities: Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Mar 1998.

Uncertainty of Measurement:

A. Type A (random) uncertainty:	$\pm 0.7\%$
B. Type B (systematic) uncertainty:	$\pm 3.0\%$
C. Uncertainty in aliquot weighing:	$\pm 0.0\%$
D. Total uncertainty at the 99% confidence level:	$\pm 3.1\%$

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 4483 α/min in 2π on 11 Apr 03.

Daniel James Van Dalsum
Quality Control

15-Apr-03
Date Signed

IPL Ref. No.: 987-7

Isotope Products Laboratories

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661•309•1010
Fax 661•257•8303

Re-Calibrated 10/4/16
New Exp Date 10/4/2017
PAI 183 JP1051
Recalibrated 4-15-05

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW040203/R2193
Catalog No.: MISC-STD
Reference Date: 1-May-03 12:00 PST
Source No.: 92MIX2203028

Contained Radioactivity:

U-234:	6.467 nCi (239.3 Bq)	Am-241:	0.6366 nCi (23.55 Bq)
U-235:	0.1135 nCi (4.200 Bq)	Total Activity:	7.217 nCi (267.1 Bq)

Physical description:

A. Capsule type:	Disk (22 mm OD X 0.79 mm THK)
B. Nature of active deposit:	Electrodeposited and diffusion bonded oxides
C. Active Diameter:	19 mm
D. Backing:	Stainless steel
E. Cover:	None

Radioimpurities: Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:

A. Type A (random) uncertainty:	$\pm 0.7\%$
B. Type B (systematic) uncertainty:	$\pm 3.0\%$
C. Uncertainty in aliquot weighing:	$\pm 0.0\%$
D. Total uncertainty at the 99% confidence level:	$\pm 3.1\%$

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 8091 α/min in 2π on 11 Apr 03.

Daniel James Van Dalsen
Quality Control

15-Apr-03
Date Signed

IPL Ref. No.: 987-7

ISO 9001 CERTIFIED

Medical Imaging Laboratory
24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street Burbank, California 91504



An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661•309•1010
Fax 661•257•8303

Re-Calibrated 10/4/16
New Exp Date 10/4/2017
JHolske

PAT I.O. 184
recalibration 4-1503

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006) \times 10^5$ years
Half Life (U-235): $(7.037 \pm 0.011) \times 10^8$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW040203/R2193
Catalog No.: MISC-STD
Reference Date: 1-May-03 12:00 PST
Source No.: 92MIX2203024

Contained Radioactivity:

U-234:	3.227 nCi (119.4 Bq)	Am-241:	2.866 nCi (106.0 Bq)
U-235:	0.05205 nCi (1.926 Bq)	Total Activity:	6.145 nCi (227.3 Bq)

Physical description:

A. Capsule type: Disk (22 mm OD X 0.79 mm THK)
B. Nature of active deposit: Electrodeposited and diffusion bonded oxides
C. Active Diameter: 19 mm
D. Backing: Stainless steel
E. Cover: None

Radioimpurities: Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:

A. Type A (random) uncertainty: $\pm 0.6\%$
B. Type B (systematic) uncertainty: $\pm 3.0\%$
C. Uncertainty in aliquot weighing: $\pm 0.0\%$
D. Total uncertainty at the 99% confidence level: $\pm 3.1\%$

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 6889 $\alpha/min \cdot ln 2\pi$ on 11 Apr 03.

Daniel James Van Dalsen
Quality Control

15-Apr-03
Date Signed

IPL Ref. No.: 987-7

ISO 9001 CERTIFIED

Medical Imaging Laboratory
24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street Burbank, California 91504

24937 Avenue Tibbitts
Valencia, California 91355Tel 661-309-1010
Fax 661-257-8303ReCalibrated 10/4/16
New Exp Date 10/4/2017
JW 10/5/16PAI ISO 00185
recd from recalibrator
3-28-03CERTIFICATE OF CALIBRATION
ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW030603/R2155
Catalog No.: MISC-STD
Reference Date: 1-Apr-03 12:00 PST
Source No.: 92MIX2203021

Contained Radioactivity:
U-234: 2.731 nCi (101.0 Bq)
U-235: 0.03416 nCi (1.264 Bq)

Am-241: 0.9325 nCi (34.50 Bq)
Total Activity: 3.698 nCi (136.8 Bq)

Physical description:

- A. Capsule type: Disk (22 mm OD X 0.79 mm THK)
B. Nature of active deposit: Electrodeposited and diffusion bonded oxides
C. Active Diameter: 19 mm
D. Backing: Stainless steel
E. Cover: None

Radioimpurities: Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:

- A. Type A (random) uncertainty: $\pm 0.8\%$
B. Type B (systematic) uncertainty: $\pm 3.1\%$
C. Uncertainty in aliquot weighing: $\pm 0.0\%$
D. Total uncertainty at the 99% confidence level: $\pm 3.2\%$

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 4145 α/min in 2π on 18 Mar 03.

Daniel James Van Dalsen
Quality Control

19-Mar-03
Date Signed

IPL Ref. No.: 987-2



Isotope Products
Laboratories
An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

ReCalibrated 10/4/16
New Exp Date 10/4/2017
TP10516

QA ID 188
feed for recalibration
3-28-03

CERTIFICATE OF CALIBRATION ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006) \times 10^5$ years
Half Life (U-235): $(7.037 \pm 0.011) \times 10^8$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW030603/R2155
Catalog No.: MISC-STD
Reference Date: 1-Apr-03 12:00 PST
Source No.: 92MIX2203023

Contained Radioactivity:

U-234:	2.895 nCi (107.1 Bq)	Am-241:	1.953 nCi (72.26 Bq)
U-235:	0.02502 nCi (0.9257 Bq)	Total Activity:	4.873 nCi (180.3 Bq)

Physical description:

A. Capsule type: Disk (22 mm OD X 0.79 mm THK)
B. Nature of active deposit: Electrodeposited and diffusion bonded oxides
C. Active Diameter: 19 mm
D. Backing: Stainless steel
E. Cover: None

Radioimpurities: Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:

A. Type A (random) uncertainty: $\pm 0.8\%$
B. Type B (systematic) uncertainty: $\pm 3.1\%$
C. Uncertainty in aliquot weighing: $\pm 0.0\%$
D. Total uncertainty at the 99% confidence level: $\pm 3.2\%$

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 5463 α/min in 2π on 18 Mar 03.

Daniel James Van Dalsen
Quality Control

19-Mar-03
Date Signed

IPL Ref. No.: 987-2

ISO 9001 CERTIFIED

Medical Imaging Laboratory
24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street Burbank, California 91504



Isotope Products
Laboratories

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

ReCalibrated 10/14/16
New Exp. Dat 10/14/2017
JP 10/15/16

PAI ID 189
recd 4-21-03
recalibrated 4-15-03

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006) \times 10^5$ years
Half Life (U-235): $(7.037 \pm 0.011) \times 10^8$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW040203/R2193
Catalog No.: MISC-STD
Reference Date: 1-May-03 12:00 PST
Source No.: 92MIX2203029

Contained Radioactivity:

U-234:	9.048 nCi	(334.8 Bq)	Am-241:	1.433 nCi	(53.02 Bq)
U-235:	0.1771 nCi	(6.553 Bq)	Total Activity:	10.66 nCi	(394.4 Bq)

Physical description:

A. Capsule type: Disk (22 mm OD X 0.79 mm THK)
B. Nature of active deposit: Electrodeposited and diffusion bonded oxides
C. Active Diameter: 19 mm
D. Backing: Stainless steel
E. Cover: None

Radioimpurities: Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Mar 1998.

Uncertainty of Measurement:

A. Type A (random) uncertainty: $\pm 0.5\%$
B. Type B (systematic) uncertainty: $\pm 3.0\%$
C. Uncertainty in aliquot weighing: $\pm 0.0\%$
D. Total uncertainty at the 99% confidence level: $\pm 3.0\%$

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 11950 α/min in 2π on 11 Apr 03.

Daniel James Van Dalsem
Quality Control

15-Apr-03
Date Signed

IPL Ref. No.: 987-7