

Session Report

4/27/2015

Information Panel

Name	Nelson Farms
Start Time	4/3/2015 3:26:05 PM
Stop Time	4/3/2015 3:57:00 PM
Device Name	BGN010014
Model Type	SoundPro DL
Device Firmware Rev	R.13F
Comments	Winds calm at time of sound study

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	58.6 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	73.6 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
4/3/2015 3:25:08 PM	Calibration	114.0			
4/3/2015 3:57:34 PM	Verification	114.2			

Logged Data Chart

Nelson Farms: Logged Data Chart

