

Session Report

3/10/2017

Information Panel


Name Carter031017
Start Time 3/10/2017 11:16:01 AM
Stop Time 3/10/2017 12:14:02 PM
Device Name BGN010009
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	43.4 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	61.7 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Logged Data Chart

Carter031017: Logged Data Chart



Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
3/10/2017 11:15:38 AM	Calibration	114.0			
3/10/2017 12:14:17 PM	Verification	113.8			